

LB 133356


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38065



A= 0.0000e+000

B= 2.6560e-004 slope

C= -2.2736e-001 y-intercept

Rho= 0.9997412

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	b/c/d
0.00	0.000	-0.026	-0.026	759	0.000	759					5
0.05	0.050					442	0				12
0.20	0.200	0.223	0.023	1697	0.0 %	1697					-2
2.50	2.500	2.442	-0.058	10050	0.0 %	10050					1
5.00	5.000	5.025	0.025	19775	0.0 %	19775					2
7.50	7.500	7.653	0.153	29669	0.0 %	29669					1
10.0	10.000	9.883	-0.117	38065	0.0 %	38065					1

LB133356

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	759	-		0 SFAM01.1	PPB	11/8/2024 13:14	Std
	0.2 S01	1697	-		0.2 SFAM01.1	PPB	11/8/2024 13:19	Std
	2.5 S02	10050	-		2.5 SFAM01.1	PPB	11/8/2024 13:22	Std
	5 S03	19775	-		5 SFAM01.1	PPB	11/8/2024 13:26	Std
	7.5 S04	29669	-		7.5 SFAM01.1	PPB	11/8/2024 13:28	Std
	10 S05	38065	-		10 SFAM01.1	PPB	11/8/2024 13:36	Std
ICV76	ICV76	17224	4.3474 -		SFAM01.1	PPB	11/8/2024 13:38	SMPL
ICB76	ICB76	686	-0.0452 -		SFAM01.1	PPB	11/8/2024 13:43	SMPL
CCV54	CCV54	19697	5.0042 -		SFAM01.1	PPB	11/8/2024 13:45	SMPL
CCB54	CCB54	813	-0.0114 -		SFAM01.1	PPB	11/8/2024 13:52	SMPL
P4496-04	MBHDO5	637	-0.0582 -		SFAM01.1	PPB	11/8/2024 13:57	SMPL
P4496-05	MBHDO6	1804	0.2518 -		SFAM01.1	PPB	11/8/2024 13:59	SMPL
P4496-06	MBHZC7	707	-0.0396 -		SFAM01.1	PPB	11/8/2024 14:01	SMPL
P4496-07	MBHZC8	841	-0.004 -		SFAM01.1	PPB	11/8/2024 14:04	SMPL
P4496-08	MBHZC8D	966	0.0292 -		SFAM01.1	PPB	11/8/2024 14:06	SMPL
P4624-01	MBH6M7	-48	-0.2401 -		SFAM01.1	PPB	11/8/2024 14:10	SMPL
P4624-02	MBH6M8	-25	-0.234 -		SFAM01.1	PPB	11/8/2024 14:13	SMPL
P4624-03	MBH6M9	248	-0.1615 -		SFAM01.1	PPB	11/8/2024 14:15	SMPL
P4624-04	MBH6N0	137	-0.191 -		SFAM01.1	PPB	11/8/2024 14:17	SMPL
P4624-05	MBH6N1	207	-0.1724 -		SFAM01.1	PPB	11/8/2024 14:19	SMPL
P4624-06	MBH8M6	138	-0.1907 -		SFAM01.1	PPB	11/8/2024 14:22	SMPL
P4624-07	MBH6N2	44	-0.2157 -		SFAM01.1	PPB	11/8/2024 14:24	SMPL
P4624-08	MBH6N3	562	-0.0781 -		SFAM01.1	PPB	11/8/2024 14:26	SMPL
P4624-09	MBH6N4	123	-0.1947 -		SFAM01.1	PPB	11/8/2024 14:28	SMPL
P4624-10	MBH6N5	79	-0.2064 -		SFAM01.1	PPB	11/8/2024 14:31	SMPL
P4624-11	MBH6N6	29	-0.2197 -		SFAM01.1	PPB	11/8/2024 14:33	SMPL
P4624-12	MBH6N6D	63	-0.2106 -		SFAM01.1	PPB	11/8/2024 14:35	SMPL
P4624-13	MBH6N6S	12093	2.9846 -		SFAM01.1	PPB	11/8/2024 14:38	SMPL
CCV55	CCV55	20269	5.1561 -		SFAM01.1	PPB	11/8/2024 14:40	SMPL
CCB55	CCB55	773	-0.0221 -		SFAM01.1	PPB	11/8/2024 14:44	SMPL
P4624-14	MBH6N7	49	-0.2143 -		SFAM01.1	PPB	11/8/2024 14:46	SMPL
PB164799BL	PBS799	815	-0.0109 -		SFAM01.1	PPB	11/8/2024 14:49	SMPL
P4501-08DLX2	MBHCT5	23118	5.9128 -		SFAM01.1	PPB	11/8/2024 14:51	SMPL
P4496-09	MBHZC8S	10873	2.6605 -		SFAM01.1	PPB	11/8/2024 14:56	SMPL
CCV56	CCV56	20110	5.1139 -		SFAM01.1	PPB	11/8/2024 14:58	SMPL
CCB56	CCB56	775	-0.0215 -		SFAM01.1	PPB	11/8/2024 15:17	SMPL