



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:50
In Date: 11/08/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 11/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB133360

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4728-01	MBH9B8	1	1.15	8.40	9.55	8.57	88.3	
P4728-02	MBH9B8D	2	1.15	8.40	9.55	8.57	88.3	
P4728-03	MBH9B8S	3	1.15	8.40	9.55	8.57	88.3	
P4728-04	MBH9B9	4	1.15	8.64	9.79	8.33	83.1	
P4728-05	MBH9C0	5	1.15	8.82	9.97	8.42	82.4	
P4728-06	MBH9C2	6	1.15	8.81	9.96	9.47	94.4	
P4728-07	MBH9C3	7	1.18	8.66	9.84	8.33	82.6	
P4728-08	MBH9C4	8	1.17	8.74	9.91	8.23	80.8	
P4728-09	MBH9D4	9	1.13	8.80	9.93	8.65	85.5	
P4728-10	MBH9D5	10	1.15	8.48	9.63	7.98	80.5	
P4728-11	MBH9D6	11	1.18	8.40	9.58	8.56	87.9	
P4728-12	MBH9F0	12	1.17	8.57	9.74	8.43	84.7	
P4728-13	MBH9F1	13	1.16	8.76	9.92	8.23	80.7	
P4728-14	MBH9F2	14	1.15	8.69	9.84	8.81	88.1	
P4728-15	MBH9G6	15	1.14	8.67	9.81	8.53	85.2	
P4728-17	MBH9F4	16	1.16	8.69	9.85	8.63	86.0	
P4728-18	MBH9F5	17	1.12	8.80	9.92	9.00	89.5	
P4728-19	MBH9F6	18	1.14	8.84	9.98	8.71	85.6	
P4728-20	MBH9F8	19	1.14	8.82	9.96	8.81	87.0	
P4728-21	MBH9F9	20	1.13	8.84	9.97	8.94	88.3	
P4728-22	MBH9G0	21	1.13	8.38	9.51	8.32	85.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$