

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133369

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/08/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	11/08/2024	16:01
3	CAL3	1	Water	NA	NA	20.2	10.02	11/08/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.00	11/08/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.02	11/08/2024	17:25
6	P4791-02	1	Water	NA	NA	20.7	6.81	11/08/2024	17:30
7	P4792-04	1	Water	NA	NA	20.2	6.86	11/08/2024	17:38
8	P4797-01	1	Water	NA	NA	22.2	7.55	11/08/2024	17:44
9	P4797-01DUP	1	Water	NA	NA	22.3	7.56	11/08/2024	17:45
10	CCV2	1	Water	NA	NA	20.3	12.02	11/08/2024	17:47

WORKLIST(Hardcopy Internal Chain)

133369

WorkList Name : ph p4797

WorkList ID : 185273

Department : Wet-Chemistry

Date : 11-08-2024 17:12:42

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4791-02	H HINCHMAN-OILY-WATER	Water	pH	Cool 4 deg C	PSEG03	L23	11/08/2024	9040C
P4792-04	H MANHOLE-WASTE-DRUMS	Water	pH	Cool 4 deg C	PSEG03	L21	11/08/2024	9040C
P4797-01	H 2885	Water	pH	Cool 4 deg C	PSEG03	L21	11/08/2024	9040C

Date/Time 11/08/24 17:12:0
Raw Sample Received by: 5860001
Raw Sample Relinquished by: 5860001

Date/Time 11/08/24 18:13:0
Raw Sample Received by: 5860001
Raw Sample Relinquished by: 5860001