

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133370

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	11/08/2024	12:40
2	P4759-01		1	14	100.00	>212.0	>212.0	11/08/2024	13:10
3	P4759-01DUP		1	14	100.00	>212.0	>212.0	11/08/2024	13:40
4	P4760-01		1	12	100.00	>212.0	>212.0	11/08/2024	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133370

WorkList Name : fp-11-08-

WorkList ID : 185249

Department : Wet-Chemistry

Date : 11-08-2024 09:13:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4759-01 E	MONTCLAIR-TOTE-1	Water	Flash Point	Cool 4 deg C	PSEG03	L23	11/07/2024	1010B
P4760-01 E	LPA-TOTE-1	Water	Flash Point	Cool 4 deg C	PSEG03	L23	11/07/2024	1010B

Date/Time 11/08/2024 12:30
Raw Sample Received by: RM CWL
Raw Sample Relinquished by: RM CWL

Date/Time 11/08/2024 14:45
Raw Sample Received by: RM CWL
Raw Sample Relinquished by: RM CWL