



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/11/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:25
In Date: 11/09/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:40
Out Date: 11/10/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133372

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4751-01	BH9C6	1	1.12	8.77	9.89	8.6	85.3	
P4751-02	BH9C7	2	1.12	8.78	9.9	8.53	84.4	
P4751-03	BH9C8	3	1.15	8.81	9.96	8.68	85.5	
P4751-04	BH9D0	4	1.15	8.43	9.58	8.43	86.4	
P4751-05	BH9D1	5	1.18	8.45	9.63	8.45	86.0	
P4751-06	BH9E3	6	1.15	8.83	9.98	8.91	87.9	
P4751-07	BH9E6	7	1.17	8.58	9.75	8.79	88.8	
P4751-08	BH9E6MS	8	1.17	8.58	9.75	8.79	88.8	
P4751-09	BH9E6MSD	9	1.17	8.58	9.75	8.79	88.8	
P4751-10	BH9E7	10	1.16	8.63	9.79	8.46	84.6	
P4751-11	BH9E8	11	1.16	8.52	9.68	8.33	84.2	
P4751-12	VHBLK001	12	1.00	1.00	2.00	2.00	100.0	vhblk
P4751-13	BH9F4	13	1.16	8.49	9.65	8.47	86.1	
P4751-14	BH9G7	14	1.19	8.63	9.82	8.92	89.6	
P4751-16	BH9F5	15	1.17	8.66	9.83	8.9	89.3	
P4751-17	BH9F6	16	1.15	8.65	9.8	8.05	79.8	
P4751-18	BH9F8	17	1.16	8.57	9.73	8.84	89.6	
P4751-19	BH9F9	18	1.16	8.53	9.69	8.52	86.3	
P4751-20	BH9G0	19	1.15	8.83	9.98	8.77	86.3	
P4751-21	BH9G2	20	1.12	8.57	9.69	8.91	90.9	
P4751-22	BH9G3	21	1.12	8.85	9.97	8.91	88.0	
P4751-23	BH9G4	22	1.16	8.49	9.65	8.46	86.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$