


LB133376

SFAM01.1

INSTRUMENT ID: LV1

Linear



A= 0.0000e+000

B= 2.4448e-004 *slope*

C= 8.2238e-003 *intercept*

Rho= 0.9997200

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.00	0.000	0.007	0.007	-6	0.000	-6					=
0.05	0.050					442	0				-17
0.20	0.200	0.167	-0.033	648	0.0 %	648					-2
2.50	2.500	2.441	-0.059	9950	0.0 %	9950					2
5.00	5.000	5.084	0.084	20762	0.0 %	20762					2
7.50	7.500	7.631	0.131	31180	0.0 %	31180					2
10.0	10.000	9.870	-0.130	40339	0.0 %	40339					-1

LB133376

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-6	-		0 SFAM01.1	PPB	11/10/2024 11:23	Std
	0.2 S01	648	-		0.2 SFAM01.1	PPB	11/10/2024 11:25	Std
	2.5 S02	9950	-		2.5 SFAM01.1	PPB	11/10/2024 11:27	Std
	5 S03	20762	-		5 SFAM01.1	PPB	11/10/2024 11:30	Std
	7.5 S04	31180	-		7.5 SFAM01.1	PPB	11/10/2024 11:32	Std
	10 S05	40339	-		10 SFAM01.1	PPB	11/10/2024 11:34	Std
ICV078	ICV078	15014	3.6789 -		SFAM01.1	PPB	11/10/2024 11:37	SMPL
ICB078	ICB078	-201	-0.0409 -		SFAM01.1	PPB	11/10/2024 11:39	SMPL
CCV059	CCV059	21289	5.213 -		SFAM01.1	PPB	11/10/2024 11:43	SMPL
CCB059	CCB059	-227	-0.0473 -		SFAM01.1	PPB	11/10/2024 11:46	SMPL
PB164851BL	PBW851	-76	-0.0104 -		SFAM01.1	PPB	11/10/2024 11:48	SMPL
P4730-01	MCOB82	-5	0.007 -		SFAM01.1	PPB	11/10/2024 11:50	SMPL
P4730-02	MCOB83	-83	-0.0121 -		SFAM01.1	PPB	11/10/2024 11:52	SMPL
P4730-03	MCOB84	-85	-0.0126 -		SFAM01.1	PPB	11/10/2024 11:55	SMPL
P4730-04	MCOB86	-72	-0.0094 -		SFAM01.1	PPB	11/10/2024 11:57	SMPL
P4730-05	MCOB86D	-81	-0.0116 -		SFAM01.1	PPB	11/10/2024 11:59	SMPL
P4730-06	MCOB86S	3354	0.8282 -		SFAM01.1	PPB	11/10/2024 12:02	SMPL
P4730-07	MCOB88	-135	-0.0248 -		SFAM01.1	PPB	11/10/2024 12:04	SMPL
CCV060	CCV060	20987	5.1392 -		SFAM01.1	PPB	11/10/2024 12:06	SMPL
CCB060	CCB060	-208	-0.0426 -		SFAM01.1	PPB	11/10/2024 12:08	SMPL