

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133385

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/11/2024	08:35
2	P4791-02		1	11	100.00	>212.0	>212.0	11/11/2024	09:05
3	P4791-02DUP		1	11	100.00	>212.0	>212.0	11/11/2024	09:35
4	P4792-04		1	13	100.00	>212.0	>212.0	11/11/2024	10:05
5	P4797-01		1	14	100.00	>212.0	>212.0	11/11/2024	10:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133385

WorkList Name : flash- point

WorkList ID : 185293

Department : Wet-Chemistry

Date : 11-09-2024 08:14:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4791-02	HINCHMAN-OILY-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	L23	11/08/2024	1010B
P4792-04	MANHOLE-WASTE-DRUMS	Water	Flash Point	Cool 4 deg C	PSEG03	L21	11/08/2024	1010B
P4797-01	2885	Water	Flash Point	Cool 4 deg C	PSEG03	L21	11/08/2024	1010B

Date/Time 11/11/2024 08:25
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 11/11/2024 11:15
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC