

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133400

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/11/2024	16:35
2	CAL2	1	Water	NA	NA	20.2	7.00	11/11/2024	16:36
3	CAL3	1	Water	NA	NA	20.3	10.02	11/11/2024	16:37
4	ICV	1	Water	NA	NA	20.2	7.00	11/11/2024	16:39
5	CCV1	1	Water	NA	NA	20.3	2.01	11/11/2024	16:40
6	P4812-03	1	Water	NA	NA	23.5	7.53	11/11/2024	16:47
7	P4812-03DUP	1	Water	NA	NA	23.6	7.54	11/11/2024	16:48
8	P4813-01	1	Water	NA	NA	23.2	6.70	11/11/2024	17:00
9	P4813-04	1	Water	NA	NA	24.6	7.72	11/11/2024	17:10
10	CCV2	1	Water	NA	NA	20.3	12.02	11/11/2024	17:15

WORKLIST(Hardcopy Internal Chain)

B133400

WorkList Name : ph p4807 water WorkList ID : 185339 Department : Wet-Chemistry Date : 11-11-2024 16:28:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4812-03	270	Water	pH	Cool 4 deg C	PSEG03	L41	11/11/2024	9040C
P4813-01	416	Water	pH	Cool 4 deg C	PSEG03	L31	11/11/2024	9040C
P4813-04	419	Water	pH	Cool 4 deg C	PSEG03	L31	11/11/2024	9040C

Date/Time 11/11/24 16:30
 Raw Sample Received by: SO wely
 Raw Sample Relinquished by: RJ cest-lob

Date/Time 11/11/24 18:30
 Raw Sample Received by: RJ cest-lob
 Raw Sample Relinquished by: SO wely