

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133406

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	11/12/2024	08:30
2	P4812-03		1	14	100.00	>212.0	>212.0	11/12/2024	09:00
3	P4812-03DUP		1	14	100.00	>212.0	>212.0	11/12/2024	09:30
4	P4813-01		1	12	100.00	>212.0	>212.0	11/12/2024	10:00
5	P4813-04		1	14	100.00	>212.0	>212.0	11/12/2024	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133406

WorkList Name : FP-11-11

WorkList ID : 185347

Department : Wet-Chemistry

Date : 11-11-2024 17:44:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4812-03	270	Water	Flash Point	Cool 4 deg C	PSEG03	L41	11/11/2024	1010B
P4813-01	416	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/11/2024	1010B
P4813-04	419	Water	Flash Point	Cool 4 deg C	PSEG03	L31	11/11/2024	1010B

Date/Time 11/12/2024 08.20
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

Date/Time 11/12/2024 11.10
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC