

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133423

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/13/2024	08:45
2	CAL2	1	Water	NA	NA	20.3	7.00	11/13/2024	08:46
3	CAL3	1	Water	NA	NA	20.3	10.02	11/13/2024	08:52
4	ICV	1	Water	NA	NA	20.3	7.00	11/13/2024	09:00
5	CCV1	1	Water	NA	NA	20.2	2.01	11/13/2024	09:10
6	P4820-01	1	Solid	20.02	20	24.6	5.44	11/13/2024	09:20
7	P4820-01DUP	1	Solid	20.03	20	24.7	5.45	11/13/2024	09:21
8	CCV2	1	Water	NA	NA	20.3	12.02	11/13/2024	09:25

WORKLIST(Hardcopy Internal Chain)

133423

WorkList Name : ph s p4820

WorkList ID : 185388

Department : Wet-Chemistry

Date : 11-13-2024 08:31:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4820-01	C	CO-SIAC-20	Solid	pH	Cool 4 deg C	PSEG03 L21	11/12/2024	9045D

Date/Time 11-13-24 08:35
Raw Sample Received by: JQ Cooper
Raw Sample Relinquished by: JQ Cooper

Date/Time 11-13-24 13:30
Raw Sample Received by: JQ Cooper
Raw Sample Relinquished by: JQ Cooper