

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133444

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/14/2024	08:15
2	CAL2	1	Water	NA	NA	20.2	7.01	11/14/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	11/14/2024	08:19
4	ICV	1	Water	NA	NA	20.3	7.00	11/14/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	11/14/2024	08:22
6	P4834-01	1	Water	NA	NA	20.7	7.09	11/14/2024	08:25
7	P4834-01DUP	1	Water	NA	NA	20.8	7.10	11/14/2024	08:26
8	P4835-01	1	Water	NA	NA	20.4	5.53	11/14/2024	08:30
9	P4836-01	1	Water	NA	NA	20.6	6.74	11/14/2024	08:33
10	P4837-01	1	Water	NA	NA	20.3	7.56	11/14/2024	08:37
11	P4838-01	1	Water	NA	NA	20.6	7.39	11/14/2024	08:40
12	P4843-01	1	Water	NA	NA	20.2	12.02	11/14/2024	08:47
13	CCV2	1	Water	NA	NA	20.3	12.01	11/14/2024	08:50

WORKLIST(Hardcopy Internal Chain)

133444

WorkList Name : ph p4838

WorkList ID : 185416

Department : Wet-Chemistry

Date : 11-14-2024 08:00:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4834-01	SLP-1-WATER	Water	pH	Cool 4 deg C	PSEG03	L41	11/13/2024	9040C
P4835-01	345-2-WATER	Water	pH	Cool 4 deg C	PSEG03	L41	11/13/2024	9040C
P4836-01	T3-WATER	Water	pH	Cool 4 deg C	PSEG03	L41	11/13/2024	9040C
P4837-01	3TRK-WATER	Water	pH	Cool 4 deg C	PSEG03	L41	11/13/2024	9040C
P4838-01	T1-WATER	Water	pH	Cool 4 deg C	PSEG03	L41	11/13/2024	9040C
P4843-01	SW-WTS-01	Water	pH	Cool 4 deg C	ENTA05	L23	11/13/2024	9040C

Date/Time

11-14-24 08:10

Raw Sample Received by:

18 Cell 1

Raw Sample Relinquished by:

Cell 5R

Date/Time

11-15-24

Raw Sample Received by:

Cell 8R

Raw Sample Relinquished by:

Cell 40C