

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133445

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	11/14/2024	11:20
2	CAL2	1	Water	NA	NA	20.2	7.00	11/14/2024	11:22
3	CAL3	1	Water	NA	NA	20.3	10.02	11/14/2024	11:25
4	ICV	1	Water	NA	NA	20.3	7.00	11/14/2024	11:30
5	CCV1	1	Water	NA	NA	20.2	2.01	11/14/2024	11:33
6	P4826-01	1	Solid	20.02	20	24.2	8.71	11/14/2024	11:40
7	P4826-01DUP	1	Solid	20.03	20	24.3	8.72	11/14/2024	11:41
8	CCV2	1	Water	NA	NA	20.3	12.02	11/14/2024	11:45

WORKLIST(Hardcopy Internal Chain)

133445

WorkList Name : ph p4826 WorkList ID : 185435 Department : Wet-Chemistry Date : 11-14-2024 11:02:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4826-01	GAS-BUR-2005	Solid	pH	Cool 4 deg C	PSEG03	L23	11/13/2024	9045D

Date/Time 11-14-24 11:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11-24-24 12:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]