

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/15/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 11/14/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 11/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133451

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P4839-01	EX-9-TPH-9	1	1.14	8.58	9.72	8.44	85.1	
P4839-02	EX-9-TPH-10	2	1.15	8.54	9.69	8.61	87.4	
P4839-03	EX-9-TPH-11	3	1.18	8.53	9.71	8.55	86.4	
P4839-04	EX-9-TPH-12	4	1.15	8.82	9.97	8.85	87.3	
P4839-05	EX-9-TPH-13	5	1.12	8.70	9.82	8.64	86.4	
P4839-06	EX-9-TPH-14	6	1.15	8.82	9.97	8.9	87.9	
P4839-07	EX-9-TPH-15	7	1.15	8.40	9.55	8.1	82.7	
P4839-08	EX-9-TPH-16	8	1.15	8.83	9.98	8.54	83.7	
P4839-09	EX-9-TPH-17	9	1.15	8.82	9.97	9.06	89.7	
P4839-10	EX-9-TPH-18	10	1.12	8.87	9.99	8.13	79.0	
P4839-11	EX-9-TPH-19	11	1.16	8.73	9.89	8.43	83.3	
P4839-12	EX-9-TPH-20	12	1.14	8.80	9.94	8.11	79.2	
P4839-13	EX-9-TPH-21	13	1.16	8.47	9.63	7.99	80.6	
P4839-14	EX-10-TPH-1	14	1.14	8.56	9.7	8.41	84.9	
P4839-15	EX-10-TPH-2	15	1.18	8.80	9.98	8.57	84.0	
P4839-16	EX-10-TPH-3	16	1.17	8.60	9.77	7.84	77.6	
P4839-17	EX-4-TPH-1	17	1.13	8.80	9.93	8.51	83.9	
P4839-18	EX-4-TPH-2	18	1.13	8.81	9.94	8.69	85.8	
P4839-19	EX-4-TPH-3	19	1.15	8.82	9.97	8.58	84.2	
P4839-20	EX-4-TPH-4	20	1.17	8.40	9.57	8.22	83.9	
P4839-21	EX-4-TPH-5	21	1.18	8.81	9.99	8.65	84.8	
P4839-22	EX-4-TPH-6	22	1.15	8.82	9.97	9.04	89.5	
P4848-01	TP-1	26	1.17	8.58	9.75	8.49	85.3	
P4849-01	RR-1	27	1.18	8.64	9.82	8.88	89.1	
P4850-01	OR-03-11142024	28	1.15	8.53	9.68	9.11	93.3	
P4850-02	OR-03-11142024-E2	29	1.18	8.60	9.78	8.81	88.7	
P4852-01	COMP-1	23	1.16	8.45	9.61	8.19	83.2	
P4852-02	COMP-2	24	1.17	8.60	9.77	8.17	81.4	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/15/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 11/14/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 11/15/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133451

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4852-03	COMP-3	25	1.15	8.81	9.96	8.94	88.4	
P4857-01	ETGI-328	30	1.16	8.50	9.66	8.75	89.3	
P4857-02	ETGI-328-E2	31	1.00	1.00	2.00	2.00	100.0	debris
P4857-03	VNJ-226	32	1.18	8.72	9.9	9.37	93.9	
P4857-04	VNJ-226-E2	33	1.15	8.76	9.91	9.38	93.9	
P4858-01	BC258780-1-1	34	1.00	1.00	2.00	2.00	100.0	pilc
P4858-02	BC258780-1-2	35	1.00	1.00	2.00	2.00	100.0	pilc
P4858-03	BC258780-2-1	36	1.00	1.00	2.00	2.00	100.0	pilc
P4858-04	BC258780-2-2	37	1.00	1.00	2.00	2.00	100.0	pilc
P4858-05	HIH269B-1-1	38	1.00	1.00	2.00	2.00	100.0	pilc
P4858-06	HIH269B-1-2	39	1.00	1.00	2.00	2.00	100.0	pilc
P4858-07	BC225734-1-1	40	1.00	1.00	2.00	2.00	100.0	pilc
P4858-08	BC225734-1-2	41	1.00	1.00	2.00	2.00	100.0	pilc
P4858-09	BC225734-2-1	42	1.00	1.00	2.00	2.00	100.0	pilc
P4858-10	BC225734-2-2	43	1.00	1.00	2.00	2.00	100.0	pilc
P4858-11	BC248150-1-1	44	1.00	1.00	2.00	2.00	100.0	pilc
P4858-12	BC248150-1-2	45	1.00	1.00	2.00	2.00	100.0	pilc
P4858-13	Y2309-0044-1-1	46	1.00	1.00	2.00	2.00	100.0	pilc
P4858-14	Y2309-0044-1-2	47	1.00	1.00	2.00	2.00	100.0	pilc
P4858-15	BC274768-1-1	48	1.00	1.00	2.00	2.00	100.0	pilc
P4858-16	BC274768-1-2	49	1.00	1.00	2.00	2.00	100.0	pilc
P4858-17	BC274601-1-1	50	1.00	1.00	2.00	2.00	100.0	pilc
P4858-18	BC274601-1-2	51	1.00	1.00	2.00	2.00	100.0	pilc
P4859-01	NB-08-111424	52	1.18	8.54	9.72	9.03	91.9	
P4859-02	NB-08-111424-E2	53	1.19	8.52	9.71	9.04	92.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$