

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133474

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	11/15/2024	13:25
2	P4843-01		1	14	100.00	>212.0	>212.0	11/15/2024	13:55
3	P4861-01		1	12	100.00	>212.0	>212.0	11/15/2024	14:25
4	P4871-01		1	13	100.00	>212.0	>212.0	11/15/2024	14:55
5	P4871-01DUP		1	13	100.00	>212.0	>212.0	11/15/2024	15:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133474

WorkList Name : fp-11-15

WorkList ID : 185478

Department : Wet-Chemistry

Date : 11-15-2024 12:15:14

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4843-01	SW-WTS-01	Water	Flash Point	Cool 4 deg C	ENTA05	L23	11/13/2024	1010B
P4861-01	WC-11-A-202411	Water	Flash Point	Cool 4 deg C	AECO02	L41	11/13/2024	1010B
P4871-01	WC-10-A-202411	Water	Flash Point	Cool 4 deg C	AECO02	M11	11/13/2024	1010B

Date/Time 11/15/2024 13:15
Raw Sample Received by: RM cwc
Raw Sample Relinquished by: JS (cwc)

Date/Time 11/15/2024 16:05
Raw Sample Received by: JP (cwc)
Raw Sample Relinquished by: RM cwc