

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133476

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/15/2024	13:30
2	CAL2	1	Water	NA	NA	20.3	7.00	11/15/2024	13:31
3	CAL3	1	Water	NA	NA	20.3	10.02	11/15/2024	13:33
4	ICV	1	Water	NA	NA	20.3	7.00	11/15/2024	13:35
5	CCV1	1	Water	NA	NA	20.3	2.01	11/15/2024	13:36
6	P4848-02	1	Solid	20.02	20	21.6	10.70	11/15/2024	13:40
7	P4848-02DUP	1	Solid	20.03	20	21.7	10.71	11/15/2024	13:41
8	P4849-02	1	Solid	20.02	20	22.2	8.25	11/15/2024	13:45
9	P4870-13	1	Solid	20.03	20	21.7	7.81	11/15/2024	13:47
10	P4870-14	1	Solid	20.04	20	21.7	8.14	11/15/2024	13:50
11	P4870-15	1	Solid	20.03	20	22.2	8.02	11/15/2024	14:00
12	P4870-16	1	Solid	20.02	20	22.1	7.95	11/15/2024	14:10
13	P4887-02	1	Solid	20.03	20	21.3	6.94	11/15/2024	14:15
14	P4887-06	1	Solid	20.03	20	22.4	9.17	11/15/2024	14:22
15	CCV2	1	Water	NA	NA	20.3	12.02	11/15/2024	14:27

WORKLIST(Hardcopy Internal Chain)

133446

WorkList Name : corrosivity p4848

WorkList ID : 185452

Department : Wet-Chemistry

Date : 11-15-2024 08:03:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4848-02	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L41	11/14/2024	9045D
P4849-02	RR-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	L51	11/14/2024	9045D
P4870-13	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03		11/14/2024	9045D
P4870-14	MH-735	Solid	Corrosivity	Cool 4 deg C	PSEG03		11/14/2024	9045D
P4870-15	MH-736	Solid	Corrosivity	Cool 4 deg C	PSEG03		11/14/2024	9045D
P4870-16	TP-15	Solid	Corrosivity	Cool 4 deg C	PSEG03		11/14/2024	9045D
P4887-02	MH-739	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	11/15/2024	9045D
P4887-06	MH-760	Solid	Corrosivity	Cool 4 deg C	PSEG03	M11	11/15/2024	9045D

Date/Time 11-15-24 13:00
 Raw Sample Received by: SO WCC
 Raw Sample Relinquished by: K.C. (sm)

Date/Time 11-15-24 16:00
 Raw Sample Received by: K.C. (sm)
 Raw Sample Relinquished by: SO WCC