

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133478

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/15/2024	17:20
2	CAL2	1	Water	NA	NA	20.2	7.01	11/15/2024	17:21
3	CAL3	1	Water	NA	NA	20.1	10.02	11/15/2024	17:22
4	ICV	1	Water	NA	NA	20.2	7.01	11/15/2024	17:25
5	CCV1	1	Water	NA	NA	20.3	2.01	11/15/2024	17:30
6	P4860-01	1	Solid	20.02	20	23.3	8.30	11/15/2024	17:33
7	P4860-01DUP	1	Solid	20.03	20	23.4	8.31	11/15/2024	17:34
8	P4860-02	1	Solid	20.04	20	23.4	6.72	11/15/2024	17:37
9	P4860-03	1	Solid	20.02	20	23.0	7.03	11/15/2024	17:40
10	P4860-04	1	Solid	20.03	20	23.0	6.95	11/15/2024	17:44
11	P4860-05	1	Solid	20.05	20	22.7	6.91	11/15/2024	17:47
12	P4860-06	1	Solid	20.02	20	22.1	6.88	11/15/2024	17:50
13	P4860-07	1	Solid	20.01	20	23.1	7.29	11/15/2024	17:55
14	P4860-08	1	Solid	20.03	20	23.0	7.33	11/15/2024	18:00
15	P4860-09	1	Solid	20.04	20	22.7	6.73	11/15/2024	18:10
16	CCV2	1	Water	NA	NA	20.3	12.02	11/15/2024	18:15
17	P4860-10	1	Solid	20.02	20	23.2	8.41	11/15/2024	18:25
18	CCV3	1	Water	NA	NA	20.2	2.01	11/15/2024	18:26

WORKLIST(Hardcopy Internal Chain)

133478

WorkList Name : ph s p4860

WorkList ID : 185497

Department : Wet-Chemistry

Date : 11-15-2024 17:06:58

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4860-01	DUP-01	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-02	PH2-BPT-001	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-03	PH2-BPT-002	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-04	PH2-BPT-003	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-05	PH2-BPT-004	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-06	PH2-BPT-009	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-07	PH2-BPT-008	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-08	PH2-BPT-007	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-09	PH2-BPT-006	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D
P4860-10	PH2-BPT-005	Solid	pH	Cool 4 deg C	CORN02	L41	11/14/2024	9045D

Date/Time 11-15-24 17:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 11-15-24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

19130

[Signature]

[Signature]