


LA133482

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.3754e-004

C= -1.6952e-001

Rho= 0.9995120

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.022	0.022	806	0.000	806					—
0.2	0.200	0.229	0.029	1676	0.0 %	1676					15
2.5	2.500	2.435	-0.065	10963	0.0 %	10963					-3
5.0	5.000	4.901	-0.099	21344	0.0 %	21344					-2
7.5	7.500	7.727	0.227	33243	0.0 %	33243					3
10.0	10.000	9.887	-0.113	42337	0.0 %	42337					-1

LB133482 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Nov 2024 09:44:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	806	-	0.0000	7471B	PPB	18 Nov 2024 10:03:32	S	Std
0.2 - 1	10.2	1676	-	0.2000	7471B	PPB	18 Nov 2024 10:05:48	S	Std
2.5 - 1	32.5	10963	-	2.5000	7471B	PPB	18 Nov 2024 10:08:05	S	Std
5.0 - 1	58	21344	-	5.0000	7471B	PPB	18 Nov 2024 10:10:21	S	Std
7.5 - 1	59.5	33243	-	7.5000	7471B	PPB	18 Nov 2024 10:12:41	S	Std
10.0 - 1	510	42337	-	10.0000	7471B	PPB	18 Nov 2024 10:15:02	S	Std
ICV88 - 1	ICV88	16937	3.8537	-	7471B	PPB	18 Nov 2024 10:20:33	U	SMPL
ICB88 - 1	ICB88	489	-0.0534	-	7471B	PPB	18 Nov 2024 10:22:50	U	SMPL
CCV96 - 1	CCV96	21543	4.9478	-	7471B	PPB	18 Nov 2024 10:25:09	U	SMPL
CCB96 - 1	CCB96	476	-0.0565	-	7471B	PPB	18 Nov 2024 10:27:24	U	SMPL
CRA - 1	CRA	1575	0.2046	-	7471B	PPB	18 Nov 2024 10:29:44	U	SMPL
HighStd - 1	HighStd	42803	9.9979	-	7471B	PPB	18 Nov 2024 10:31:59	U	SMPL
ChkStd - 1	ChkStd	30649	7.1109	-	7471B	PPB	18 Nov 2024 10:34:14	U	SMPL
PB165042BL - 1	PBS	816	0.0243	-	7471B	PPB	18 Nov 2024 10:39:32	U	SMPL
PB165042BS - 1	LCSS	15184	3.4373	-	7471B	PPB	18 Nov 2024 10:44:09	U	SMPL
P4848-01 - 1	TP-1	23528	5.4193	-	7471B	PPB	18 Nov 2024 10:51:29	U	SMPL
P4848-01DUP - 1	TP-1DUP	22858	5.2602	-	7471B	PPB	18 Nov 2024 10:53:47	U	SMPL
P4848-01MS - 1	TP-1MS	38464	8.9672	-	7471B	PPB	18 Nov 2024 10:56:09	U	SMPL
P4848-01MSD - 1	TP-1MSD	37944	8.8437	-	7471B	PPB	18 Nov 2024 10:58:29	U	SMPL
P4849-01 - 1	RR-1	3249	0.6022	-	7471B	PPB	18 Nov 2024 11:00:50	U	SMPL
CCV97 - 1	CCV97	21295	4.8889	-	7471B	PPB	18 Nov 2024 11:03:16	U	SMPL
CCB97 - 1	CCB97	473	-0.0572	-	7471B	PPB	18 Nov 2024 11:05:33	U	SMPL
P4850-01 - 1	OR-03-11142024	3027	0.5495	-	7471B	PPB	18 Nov 2024 11:07:53	U	SMPL
P4851-01 - 1	OK-02-11142024	2269	0.3695	-	7471B	PPB	18 Nov 2024 11:10:07	U	SMPL
P4852-01 - 1	COMP-1	4311	0.8545	-	7471B	PPB	18 Nov 2024 11:12:24	U	SMPL
P4852-02 - 1	COMP-2	4575	0.9172	-	7471B	PPB	18 Nov 2024 11:14:41	U	SMPL
P4852-03 - 1	COMP-3	8026	1.7370	-	7471B	PPB	18 Nov 2024 11:16:59	U	SMPL
P4857-01 - 1	ETGI-328	445	-0.0638	-	7471B	PPB	18 Nov 2024 11:19:16	U	SMPL
P4857-03 - 1	VNJ-226	4548	0.9108	-	7471B	PPB	18 Nov 2024 11:21:35	U	SMPL
P4859-01 - 1	NB-08-111424	1355	0.1523	-	7471B	PPB	18 Nov 2024 11:23:51	U	SMPL
P4870-01 - 1	TP-1	3725	0.7153	-	7471B	PPB	18 Nov 2024 11:26:10	U	SMPL
P4870-04 - 1	MH-735	2651	0.4602	-	7471B	PPB	18 Nov 2024 11:28:26	U	SMPL
CCV98 - 1	CCV98	21986	5.0530	-	7471B	PPB	18 Nov 2024 11:30:43	U	SMPL
CCB98 - 1	CCB98	486	-0.0541	-	7471B	PPB	18 Nov 2024 11:32:59	U	SMPL
P4870-07 - 1	MH-736	2294	0.3754	-	7471B	PPB	18 Nov 2024 11:35:18	U	SMPL
P4870-10 - 1	TP-15	1416	0.1668	-	7471B	PPB	18 Nov 2024 11:37:34	U	SMPL
P4848-01LX5 - 1		4858	0.9845	-	7471B	PPB	18 Nov 2024 11:39:51	U	SMPL
P4848-01A - 1		41264	9.6324	-	7471B	PPB	18 Nov 2024 11:42:06	U	SMPL
CCV99 - 1	CCV99	22417	5.1554	-	7471B	PPB	18 Nov 2024 11:44:23	U	SMPL
CCB99 - 1	CCB99	517	-0.0467	-	7471B	PPB	18 Nov 2024 11:46:42	U	SMPL