



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/19/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:40
In Date: 11/18/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 11/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLIDS-OVEN

QC:LB133493

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4879-01	E29X1	1	1.15	8.46	9.61	8.26	84.0	
P4879-02	E29X2	2	1.12	8.57	9.69	8.28	83.5	
P4879-03	E29X3	3	1.15	8.82	9.97	8.64	84.9	
P4879-04	E29X4	4	1.12	8.85	9.97	8.82	87.0	
P4879-05	E29X5	5	1.16	8.80	9.96	7.75	74.9	
P4879-06	E29X5D	6	1.16	8.80	9.96	7.75	74.9	
P4879-07	E29X5S	7	1.16	8.80	9.96	7.75	74.9	
P4879-08	E29X6	8	1.16	8.75	9.91	8.31	81.7	
P4879-09	E29X7	9	1.14	8.83	9.97	9.34	92.9	
P4879-10	E29X8	10	1.12	8.57	9.69	8.26	83.3	
P4879-11	E29X9	11	1.12	8.85	9.97	5.62	50.8	
P4879-12	E29Y0	12	1.15	8.78	9.93	9.03	89.7	
P4879-13	E29Y1	13	1.19	8.52	9.71	7.67	76.1	
P4879-14	E29Y2	14	1.16	8.81	9.97	9.00	89.0	
P4879-15	E29Y3	15	1.12	8.82	9.94	8.97	89.0	
P4879-16	E29Y4	16	1.15	8.82	9.97	8.51	83.4	
P4879-17	E29Y5	17	1.19	8.65	9.84	8.45	83.9	
P4879-18	E29Y9	18	1.16	8.55	9.71	8.28	83.3	
P4879-19	E29Z0	19	1.13	8.79	9.92	8.37	82.4	
P4879-20	E29Y6	20	1.14	8.84	9.98	8.94	88.2	
P4879-21	E29Y7	21	1.19	8.64	9.83	7.91	77.8	
P4879-22	E2A11	22	1.13	8.84	9.97	8.14	79.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$