

LB133509

7470A

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

39112

A= 0.0000e+000

B= 2.5757e-004

C= 1.1210e-002

Rho= 0.9998704

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Sum
0.0	0.000	0.054	0.054	167	0.000	167	0				167
0.2	0.200	0.232	0.032	856	0.0 %	856					16
2.5	2.500	2.443	-0.057	9442	0.0 %	9442					2
5.0	5.000	4.952	-0.048	19183	0.0 %	19183					1
7.5	7.500	7.433	-0.067	28816	0.0 %	28816					1
10.0	10.000	10.085	0.085	39112	0.0 %	39112					1

LB133509 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Nov 2024 10:56:41

Sample ID	Extended ID	u Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	167	-	0.0000	7470A	PPB	19 Nov 2024 11:05:05	S	Std
0.2 - 1	50.2	856	-	0.2000	7470A	PPB	19 Nov 2024 11:07:21	S	Std
2.5 - 1	52.5	9442	-	2.5000	7470A	PPB	19 Nov 2024 11:09:38	S	Std
5.0 - 1	55	19183	-	5.0000	7470A	PPB	19 Nov 2024 11:11:55	S	Std
7.5 - 1	57.5	28816	-	7.5000	7470A	PPB	19 Nov 2024 11:14:12	S	Std
10.0 - 1	510	39112	-	10.0000	7470A	PPB	19 Nov 2024 11:16:29	S	Std
ICV90 - 1	ICV90	14012	3.6203	-	7470A	PPB	19 Nov 2024 11:19:21	U	SMPL
ICB90 - 1	ICB90	-130	-0.0223	-	7470A	PPB	19 Nov 2024 11:21:36	U	SMPL
CCV05 - 1	CCV05	20709	5.3452	-	7470A	PPB	19 Nov 2024 11:26:25	U	SMPL
CCB05 - 1	CCB05	-235	-0.0493	-	7470A	PPB	19 Nov 2024 11:28:40	U	SMPL
CRA - 1	CRA	630	0.1735	-	7470A	PPB	19 Nov 2024 11:30:56	U	SMPL
HighStd - 1	HighStd	38765	9.9959	-	7470A	PPB	19 Nov 2024 11:33:11	U	SMPL
ChkStd - 1	ChkStd	27345	7.0545	-	7470A	PPB	19 Nov 2024 11:35:26	U	SMPL
PB165079BL - 1	PBW	-387	-0.0885	-	7470A	PPB	19 Nov 2024 11:37:44	U	SMPL
PB165079BS - 1	LCSW	13428	3.4699	-	7470A	PPB	19 Nov 2024 11:40:02	U	SMPL
P4848-02 - 1	TP-1	-192	-0.0382	-	7470A	PPB	19 Nov 2024 11:42:18	U	SMPL
P4849-02 - 1	RR-1	-80	-0.0094	-	7470A	PPB	19 Nov 2024 11:44:35	U	SMPL
P4870-13 - 1	TP-1	6	0.0128	-	7470A	PPB	19 Nov 2024 11:46:52	U	SMPL
P4870-14 - 1	MH-735	-231	-0.0483	-	7470A	PPB	19 Nov 2024 11:49:09	U	SMPL
P4870-15 - 1	MH-736	-90	-0.0120	-	7470A	PPB	19 Nov 2024 11:51:25	U	SMPL
CCV06 - 1	CCV06	20251	5.2273	-	7470A	PPB	19 Nov 2024 11:53:40	U	SMPL
CCB06 - 1	CCB06	-168	-0.0321	-	7470A	PPB	19 Nov 2024 11:55:56	U	SMPL
P4870-16 - 1	TP-15	-60	-0.0042	-	7470A	PPB	19 Nov 2024 11:58:13	U	SMPL
P4887-02 - 1	MH-739	5	0.0125	-	7470A	PPB	19 Nov 2024 12:00:28	U	SMPL
P4887-06 - 1	MH-760	-27	0.0043	-	7470A	PPB	19 Nov 2024 12:02:44	U	SMPL
P4887-06DUP - 1	MH-760DUP	-35	0.0022	-	7470A	PPB	19 Nov 2024 12:04:59	U	SMPL
P4887-06MS - 1	MH-760MS	15604	4.0303	-	7470A	PPB	19 Nov 2024 12:07:15	U	SMPL
P4887-06MSD - 1	MH-760MSD	15791	4.0785	-	7470A	PPB	19 Nov 2024 12:09:31	U	SMPL
P4890-06 - 1	D3721	-223	-0.0462	-	7470A	PPB	19 Nov 2024 12:11:49	U	SMPL
P4892-03 - 1	WB-310-BOT	-22	0.0055	-	7470A	PPB	19 Nov 2024 12:14:07	U	SMPL
P4893-04 - 1	MH-763	-42	0.0004	-	7470A	PPB	19 Nov 2024 12:16:24	U	SMPL
P4893-08 - 1	MH-762	50	0.0241	-	7470A	PPB	19 Nov 2024 12:18:39	U	SMPL
CCV07 - 1	CCV07	19504	5.0349	-	7470A	PPB	19 Nov 2024 12:20:55	U	SMPL
CCB07 - 1	CCB07	-130	-0.0223	-	7470A	PPB	19 Nov 2024 12:23:10	U	SMPL
P4910-04 - 1	MH-COTTAGE	-7	0.0094	-	7470A	PPB	19 Nov 2024 12:25:28	U	SMPL
P4910-08 - 1	MH-759	5	0.0125	-	7470A	PPB	19 Nov 2024 12:27:43	U	SMPL
PB165080BL - 1	PBW	-87	-0.0112	-	7470A	PPB	19 Nov 2024 12:29:58	U	SMPL
PB165080BS - 1	LCSW	13493	3.4866	-	7470A	PPB	19 Nov 2024 12:38:27	U	SMPL
P4861-01 - 1	WC-11-A-202411	-180	-0.0352	-	7470A	PPB	19 Nov 2024 12:40:43	U	SMPL
P4861-01DUP - 1	WC-11-A-202411DUP	-36	0.0019	-	7470A	PPB	19 Nov 2024 12:43:00	U	SMPL
P4861-01MS - 1	WC-11-A-202411MS	17312	4.4703	-	7470A	PPB	19 Nov 2024 13:01:06	U	SMPL
P4861-01MSD - 1	WC-11-A-202411MSD	17158	4.4306	-	7470A	PPB	19 Nov 2024 13:03:31	U	SMPL
P4871-01 - 1	WC-10-A-202411	-229	-0.0478	-	7470A	PPB	19 Nov 2024 13:05:59	U	SMPL
P4890-01 - 1	#3711	-43	0.0001	-	7470A	PPB	19 Nov 2024 13:08:24	U	SMPL
CCV08 - 1	CCV08	19596	5.0586	-	7470A	PPB	19 Nov 2024 13:10:39	U	SMPL
CCB08 - 1	CCB08	-275	-0.0596	-	7470A	PPB	19 Nov 2024 13:12:54	U	SMPL
P4890-02 - 1	D3617	617	0.1701	-	7470A	PPB	19 Nov 2024 13:15:12	U	SMPL
P4890-03 - 1	D3690	0	0.0112	-	7470A	PPB	19 Nov 2024 13:17:27	U	SMPL
PB164986TB - 1	PB164986TB	-55	-0.0030	-	7470A	PPB	19 Nov 2024 13:19:44	U	SMPL
PB165019TB - 1	PB165019TB	35	0.0202	-	7470A	PPB	19 Nov 2024 13:21:59	U	SMPL
P4887-06LX5 - 1		-87	-0.0112	-	7470A	PPB	19 Nov 2024 13:24:15	U	SMPL
P4887-06A - 1		15359	3.9672	-	7470A	PPB	19 Nov 2024 13:26:31	U	SMPL
P4861-01LX5 - 1		-189	-0.0375	-	7470A	PPB	19 Nov 2024 13:28:47	U	SMPL
P4861-01A - 1		16297	4.2088	-	7470A	PPB	19 Nov 2024 13:31:05	U	SMPL
CCV09 - 1	CCV09	20666	5.3342	-	7470A	PPB	19 Nov 2024 13:33:21	U	SMPL
CCB09 - 1	CCB09	-138	-0.0243	-	7470A	PPB	19 Nov 2024 13:35:39	U	SMPL