

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/20/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 11/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 11/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133513

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
P4916-01	TP-1-WC	1	1.15	8.48	9.63	9.26	95.6	
P4916-02	TP-1-EPH	2	1.15	8.53	9.68	9.17	94.0	
P4916-03	TP-1-VOC	3	1.14	8.76	9.9	9.52	95.7	
P4916-05	TP-2-WC	4	1.17	8.80	9.97	9.63	96.1	
P4916-06	TP-2-EPH	5	1.16	8.47	9.63	9.31	96.2	
P4916-07	TP-2-VOC	6	1.18	8.41	9.59	9.27	96.2	
P4916-09	TP-3-WC	7	1.14	8.66	9.8	8.91	89.7	
P4916-10	TP-3-EPH	8	1.15	8.82	9.97	8.68	85.4	
P4916-11	TP-3-VOC	9	1.14	8.63	9.77	8.68	87.4	
P4917-01	MX1007	10	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4917-02	MX1008	11	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4918-01	ROOF-1-CHIMNEY-GREY	12	1.00	1.00	2.00	2.00	100.0	caluk
P4918-02	ROOF-1-2A-2B-2C	13	1.00	1.00	2.00	2.00	100.0	caluk
P4918-03	ROOF-2-3A-3B-3C	14	1.00	1.00	2.00	2.00	100.0	caluk
P4918-04	EXTREIOR-EXIT-2	15	1.00	1.00	2.00	2.00	100.0	caluk
P4918-05	ROOM-126	16	1.00	1.00	2.00	2.00	100.0	caluk
P4919-01	ROOF-1-OFF-WHITE	17	1.00	1.00	2.00	2.00	100.0	caluk
P4919-02	ROOF-1-BLACK	18	1.00	1.00	2.00	2.00	100.0	caluk
P4919-03	ROOF-11	19	1.00	1.00	2.00	2.00	100.0	caluk
P4922-01	W029060-END-1-1	20	1.00	1.00	2.00	2.00	100.0	pilc
P4922-02	W029060-END-1-2	21	1.00	1.00	2.00	2.00	100.0	pilc
P4923-02	COMP-1	22	1.00	1.00	2.00	2.00	100.0	debris
P4923-03	COMP-2	23	1.00	1.00	2.00	2.00	100.0	debris
P4923-04	COMP-3	24	1.00	1.00	2.00	2.00	100.0	debris
P4923-05	COMP-4	25	1.00	1.00	2.00	2.00	100.0	debris
P4923-06	72-11991	26	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
P4924-01	MH-4	27	1.18	8.58	9.76	9.26	94.2	
P4924-02	MH-4-EPH	28	1.15	8.52	9.67	9.22	94.7	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/20/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 16:45
In Date: 11/19/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:14
Out Date: 11/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133513

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4924-03	MH-4-VOC	29	1.12	8.57	9.69	9.43	97.0	
P4925-01	MH-741	30	1.19	8.42	9.61	8.6	88.0	
P4925-02	MH-741-EPH	31	1.19	8.55	9.74	8.59	86.5	
P4925-03	MH-741-VOC	32	1.16	8.47	9.63	8.66	88.5	
P4925-05	MH-758	33	1.14	8.59	9.73	8.86	89.9	
P4925-06	MH-758-EPH	34	1.18	8.43	9.61	8.74	89.7	
P4925-07	MH-758-VOC	35	1.12	8.70	9.82	8.89	89.3	
P4926-01	111524-A	36	1.00	1.00	2.00	2.00	100.0	wipe sample
P4926-02	111524-B	37	1.00	1.00	2.00	2.00	100.0	wipe sample
P4928-01	BC215909-1-1	38	1.00	1.00	2.00	2.00	100.0	oilc
P4928-02	BC215909-1-2	39	1.00	1.00	2.00	2.00	100.0	oilc
P4929-01	ARS520	40	1.15	8.80	9.95	8.97	88.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$