

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133519

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	11/19/2024	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	11/19/2024	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	11/19/2024	16:02
4	ICV	1	Water	NA	NA	20.3	7.00	11/19/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.01	11/19/2024	16:11
6	P4909-01	1	Solid	20.03	20	24.4	7.91	11/19/2024	17:00
7	P4909-01DUP	1	Solid	20.04	20	24.5	7.92	11/19/2024	17:01
8	CCV2	1	Water	NA	NA	20.3	12.02	11/19/2024	17:10

WORKLIST(Hardcopy Internal Chain)

VB 133519

WorkList Name : ph p4909 s

WorkList ID : 185584

Department : Wet-Chemistry

Date : 11-19-2024 15:41:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4909-01	BU-02-111824	Solid	pH	Cool 4 deg C	PSEG05	L51	11/18/2024	9045D

Date/Time 11-19-24 15:50

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time

11-19-24

Raw Sample Received by:

[Signature]

Raw Sample Relinquished by:

[Signature]