

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133526

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/20/2024	08:10
2	CAL2	1	Water	NA	NA	20.1	7.00	11/20/2024	08:11
3	CAL3	1	Water	NA	NA	20.2	10.02	11/20/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.00	11/20/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.02	11/20/2024	08:20
6	P4921-01	1	Water	NA	NA	20.7	7.02	11/20/2024	08:30
7	P4921-01DUP	1	Water	NA	NA	20.8	7.03	11/20/2024	08:31
8	P4927-01	1	Water	NA	NA	21.6	5.55	11/20/2024	08:35
9	CCV2	1	Water	NA	NA	20.3	12.02	11/20/2024	08:40

LB133526

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P4921 WATER

WorkList ID : 185590

Department : Wet-Chemistry

Date : 11-20-2024 07:52:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4921-01 I	WC-11-A-202411	Water	pH	Cool 4 deg C	AECO02	L61	11/19/2024	9040C
P4927-01 A	111424-C	Water	pH	Cool 4 deg C	PSEG03	L61	11/19/2024	9040C

Date/Time 11-20-24 08:00

Raw Sample Received by: JD WCC

Raw Sample Relinquished by: AP SM

Date/Time 11-20-24 11:00

Raw Sample Received by: AP SM

Raw Sample Relinquished by: JP WCC