

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133527

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/20/2024	13:00
2	P4921-01		1	11	100.00	>212.0	>212.0	11/20/2024	13:30
3	P4927-01		1	13	100.00	>212.0	>212.0	11/20/2024	14:00
4	P4927-01DUP		1	13	100.00	>212.0	>212.0	11/20/2024	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133527

WorkList Name : FP 11-19

WorkList ID : 185585

Department : Wet-Chemistry

Date : 11-19-2024 16:50:32

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4921-01	WC-11-A-202411	Water	Flash Point	Cool 4 deg C	AECO02	L61	11/19/2024	1010B
P4927-01	111424-C	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/19/2024	1010B

Date/Time 11/20/2024 12:50
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC 1

Date/Time 11/20/2024 15:05
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC