

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133528

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	11/21/2024	11:45
2	P4923-09		1	10	59.00	138.2	138.5	11/21/2024	12:15
3	P4923-09DUP		1	10	59.00	138.2	138.5	11/21/2024	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133528

WorkList Name : fp-11-20

WorkList ID : 185608

Department : Wet-Chemistry

Date : 11-20-2024 10:32:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4923-09	SOM-24-00118	Solid	Flash Point	Cool 4 deg C	PSEG03	L61	11/19/2024	1010B

Date/Time 11/21/2024 11:35
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

Date/Time 11/21/2024 13:20
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC