



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/21/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:25
In Date: 11/20/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 11/21/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133532

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4932-01	MBHHQ1	1	1.15	8.37	9.52	8.04	82.3	
P4932-02	MBHHQ2	2	1.15	8.77	9.92	8.55	84.4	
P4932-03	MBHHQ2D	3	1.15	8.77	9.92	8.55	84.4	
P4932-04	MBHHQ2S	4	1.15	8.77	9.92	8.55	84.4	
P4932-05	MBHHQ3	5	1.14	8.80	9.94	8.53	84.0	
P4932-06	MBHHQ4	6	1.19	8.66	9.85	8.11	79.9	
P4932-07	MBHHQ5	7	1.13	8.76	9.89	8.37	82.6	
P4932-08	MBHHQ6	8	1.17	8.56	9.73	8.38	84.2	
P4932-09	MBHHQ7	9	1.16	8.73	9.89	8.92	88.9	
P4932-10	MBHHQ8	10	1.13	8.81	9.94	7.98	77.8	
P4932-11	MBHHQ9	11	1.13	8.68	9.81	8.56	85.6	
P4932-12	MBHHP1	12	1.19	8.62	9.81	7.58	74.1	
P4932-13	MBHHP2	13	1.15	8.82	9.97	8.28	80.8	
P4932-14	MBHHP3	14	1.13	8.80	9.93	8.39	82.5	
P4932-15	MBHHP4	15	1.18	8.76	9.94	8.7	85.8	
P4932-16	MBHHP5	16	1.19	8.53	9.72	8.44	85.0	
P4932-17	MBHHP6	17	1.16	8.52	9.68	8.2	82.6	
P4932-18	MBHHP7	18	1.16	8.54	9.7	7.37	72.7	
P4932-19	MBHHP8	19	1.17	8.79	9.96	7.56	72.7	
P4932-20	MBHHP9	20	1.19	8.71	9.9	7.96	77.7	
P4932-21	MBHHQ0	21	1.13	8.70	9.83	8.02	79.2	
P4932-22	MBHHR0	22	1.15	8.81	9.96	7.74	74.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$