

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB133536

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP110780
Residual chlorine ICV-LCS, 0.4PPM	WP110779
Chlorine Calibration std, 0.1ppm	WP110774
Chlorine Calibration std, 0.2ppm	WP110775
Chlorine Calibration std, 0.8ppm	WP110776
Chlorine Calibration std, 0.0ppm	WP110773
Chlorine Calibration std, 1.6ppm	WP110777
Residual Chlorine Calibration and CCV std, 0	WP110778
Total Chlorine Powder Pillows	W3147

Intercept: 0.0043

Slope: 0.9949

Regression: 0.999878

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		11/20/2024	14:00
2	CAL2	0.1	1	0.000	0.100	0.100	0.10	-4	11/20/2024	14:03
3	CAL3	0.2	1	0.000	0.220	0.220	0.22	8.5	11/20/2024	14:06
4	CAL4	0.4	1	0.000	0.400	0.400	0.40	-0.5	11/20/2024	14:09
5	CAL5	0.8	1	0.000	0.790	0.790	0.79	-1.3	11/20/2024	14:12
6	CAL6	1.6	1	0.000	1.600	1.600	1.60	0.3	11/20/2024	14:15

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Niha

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB133536

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4080	11/20/2024	14:18
2	ICB				1	0.0000	0.0000	0.0000	-0.0040	11/20/2024	14:21
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.3980	11/20/2024	14:24
4	CCB1				1	0.0000	0.0000	0.0000	-0.0040	11/20/2024	14:27
5	LB133536BL	50	50		1	0.0000	0.0000	0.0000	-0.0040	11/20/2024	14:30
6	LB133536BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.4080	11/20/2024	14:33
7	P4368-09	50	50		1	0.0000	0.0500	0.0500	0.0460	11/20/2024	14:36
8	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3980	11/20/2024	14:39
9	CCB2				1	0.0000	0.0000	0.0000	-0.0040	11/20/2024	14:42