


LAB133542

245.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.014	-0.014	15	0.000	15					
0.05	0.050	0.037	-0.013	230	0.0 %	230	0				~26
0.2	0.200	0.211	0.011	961	0.0 %	961					5
2.5	2.500	2.517	0.017	10633	0.0 %	10633					1
5.0	5.000	4.934	-0.066	20767	0.0 %	20767					1
7.5	7.500	7.662	0.162	32209	0.0 %	32209					2
10.0	10.000	9.903	-0.097	41610	0.0 %	41610					1

LB133542 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Nov 2024 17:07:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	15	-	0.0000	245.1	PPB	20 Nov 2024 17:15:41	S	Std
0.05 - 1	50.05	230	-	0.0500	245.1	PPB	20 Nov 2024 17:20:49	S	Std
0.2 - 1	50.2	961	-	0.2000	245.1	PPB	20 Nov 2024 17:23:06	S	Std
2.5 - 1	52.5	10633	-	2.5000	245.1	PPB	20 Nov 2024 17:25:22	S	Std
5.0 - 1	55.0	20767	-	5.0000	245.1	PPB	20 Nov 2024 17:27:38	S	Std
7.5 - 1	57.5	32209	-	7.5000	245.1	PPB	20 Nov 2024 17:29:54	S	Std
10.0 - 1	51.0	41610	-	10.0000	245.1	PPB	20 Nov 2024 17:32:12	S	Std
ICV93 - 1	ICV93	16510	3.9186	-245.1	PPB	PPB	20 Nov 2024 17:40:35	U	SMPL
ICB93 - 1	ICB93	-222	-0.0708	-245.1	PPB	PPB	20 Nov 2024 17:42:49	U	SMPL
CCV18 - 1	CCV18	20442	4.8562	-245.1	PPB	PPB	20 Nov 2024 17:45:06	U	SMPL
CCB18 - 1	CCB18	-200	-0.0656	-245.1	PPB	PPB	20 Nov 2024 17:47:21	U	SMPL
CRA - 1	CRA	245	0.0405	-245.1	PPB	PPB	20 Nov 2024 17:52:43	U	SMPL
HighStd - 1	HighStd	41644	9.9114	-245.1	PPB	PPB	20 Nov 2024 17:54:57	U	SMPL
ChkStd - 1	ChkStd	28899	6.8726	-245.1	PPB	PPB	20 Nov 2024 17:57:12	U	SMPL
PB165146BL - 1	PBW	-76	-0.0360	-245.1	PPB	PPB	20 Nov 2024 18:02:11	U	SMPL
PB165146BS - 1	LCSW	16199	3.8445	-245.1	PPB	PPB	20 Nov 2024 18:04:27	U	SMPL
P4762-02 - 1	EFF-WW	-267	-0.0815	-245.1	PPB	PPB	20 Nov 2024 18:06:43	U	SMPL
P4762-02DUP - 1	EFF-WWDUP	-83	-0.0377	-245.1	PPB	PPB	20 Nov 2024 18:09:01	U	SMPL
P4762-02MS - 1	EFF-WWMS	13976	3.3145	-245.1	PPB	PPB	20 Nov 2024 18:11:18	U	SMPL
P4762-02MSD - 1	EFF-WWMSD	14926	3.5410	-245.1	PPB	PPB	20 Nov 2024 18:13:35	U	SMPL
CCV19 - 1	CCV19	20139	4.7839	-245.1	PPB	PPB	20 Nov 2024 18:15:51	U	SMPL
CCB19 - 1	CCB19	-240	-0.0751	-245.1	PPB	PPB	20 Nov 2024 18:18:08	U	SMPL
P4762-02LX5 - 1		-226	-0.0718	-245.1	PPB	PPB	20 Nov 2024 18:20:25	U	SMPL
P4762-02A - 1		14438	3.4246	-245.1	PPB	PPB	20 Nov 2024 18:22:40	U	SMPL
CCV20 - 1	CCV20	19997	4.7500	-245.1	PPB	PPB	20 Nov 2024 18:24:56	U	SMPL
CCB20 - 1	CCB20	-246	-0.0765	-245.1	PPB	PPB	20 Nov 2024 18:27:12	U	SMPL