


LB133554

SFAM01.1



Linear

A= 0.0000e+000

B= 2.6394e-004 *slope*

C= 1.5964e-002 *intercept*

Rho= 0.9987100

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>o/p D</i>
0.00	0.000	0.020	0.020	16	0.000	16					1
0.05	0.050					442	0				16
0.20	0.200	0.231	0.031	815	0.0 %	815					3
2.50	2.500	2.437	-0.063	9173	0.0 %	9173					3
5.00	5.000	4.849	-0.151	18311	0.0 %	18311					3
7.50	7.500	7.872	0.372	29765	0.0 %	29765					5
10.0	10.000	9.790	-0.210	37032	0.0 %	37032					2

LB133554

NSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	16	-		0 SFAM01.1	PPB	11/21/2024 15:14	Std
	0.2 S01	815	-		0.2 SFAM01.1	PPB	11/21/2024 15:17	Std
	2.5 S02	9173	-		2.5 SFAM01.1	PPB	11/21/2024 15:21	Std
	5 S03	18311	-		5 SFAM01.1	PPB	11/21/2024 15:23	Std
	7.5 S04	29765	-		7.5 SFAM01.1	PPB	11/21/2024 15:28	Std
	10 S05	37032	-		10 SFAM01.1	PPB	11/21/2024 15:30	Std
ICV095	ICV095	15100	4.0015 -		SFAM01.1	PPB	11/21/2024 15:36	SMPL
ICB095	ICB095	-147	-0.0228 -		SFAM01.1	PPB	11/21/2024 15:38	SMPL
CCV025	CCV025	17475	4.6284 -		SFAM01.1	PPB	11/21/2024 15:40	SMPL
CCB025	CCB025	-189	-0.0339 -		SFAM01.1	PPB	11/21/2024 15:42	SMPL
PB165174BL	PBW174	-73	-0.0033 -		SFAM01.1	PPB	11/21/2024 15:45	SMPL
P4657-01	MC0VD6	-113	-0.0139 -		SFAM01.1	PPB	11/21/2024 15:47	SMPL
P4657-02	MC0VH5	-106	-0.012 -		SFAM01.1	PPB	11/21/2024 15:49	SMPL
P4657-03	MC0VM8	-65	-0.0012 -		SFAM01.1	PPB	11/21/2024 15:51	SMPL
P4728-16	MBH9G9	45	0.0278 -		SFAM01.1	PPB	11/21/2024 15:54	SMPL
P4750-22	MBH9H0	-13	0.0125 -		SFAM01.1	PPB	11/21/2024 15:56	SMPL
CCV026	CCV026	17072	4.522 -		SFAM01.1	PPB	11/21/2024 15:58	SMPL
CCB026	CCB026	-132	-0.0189 -		SFAM01.1	PPB	11/21/2024 16:01	SMPL