

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133571

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/22/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	11/22/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	11/22/2024	08:14
4	ICV	1	Water	NA	NA	20.3	7.01	11/22/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	11/22/2024	08:22
6	P4947-01	1	Water	NA	NA	20.6	9.06	11/22/2024	08:30
7	P4947-01DUP	1	Water	NA	NA	20.7	9.07	11/22/2024	08:31
8	P4949-01	1	Water	NA	NA	20.5	7.22	11/22/2024	08:37
9	P4950-01	1	Water	NA	NA	22.4	7.54	11/22/2024	08:40
10	P4952-02	1	Water	NA	NA	22.6	5.56	11/22/2024	08:47
11	P4952-03	1	Water	NA	NA	22.6	5.80	11/22/2024	08:55
12	P4953-01	1	Water	NA	NA	22.6	7.92	11/22/2024	09:00
13	CCV2	1	Water	NA	NA	20.3	12.02	11/22/2024	09:05

WORKLIST(Hardcopy Internal Chain)

VB 133571

WorkList Name : ph p4947

WorkList ID : 185667

Department : Wet-Chemistry

Date : 11-22-2024 07:49:03

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4947-01	A A3988	Water	pH	Cool 4 deg C	PSEG03	L61	11/21/2024	9040C
P4949-01	A REACTOR-SUMP	Water	pH	Cool 4 deg C	PSEG03	L61	11/21/2024	9040C
P4950-01	A AUD-1356	Water	pH	Cool 4 deg C	PSEG03	L61	11/21/2024	9040C
P4952-02	A GAS-BUR-1617	Water	pH	Cool 4 deg C	PSEG03	L61	11/21/2024	9040C
P4952-03	A GAS-BUR-1618	Water	pH	Cool 4 deg C	PSEG03	L61	11/21/2024	9040C
P4953-01	A 1297	Water	pH	Cool 4 deg C	PSEG03	L61	11/21/2024	9040C

Date/Time 11-22-24 07:52
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM

Date/Time 11-22-24 11:00
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JDCSM