



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/25/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:05
In Date: 11/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:42
Out Date: 11/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133573

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4940-01	MBHHS7	1	1.15	8.63	9.78	7.38	72.2	
P4940-02	MBHHS8	2	1.18	8.63	9.81	8.86	89.0	
P4940-03	MBHHS9	3	1.15	8.54	9.69	8.13	81.7	
P4940-04	MBHHT0	4	1.15	8.82	9.97	8.2	79.9	
P4940-05	MBHHT1	5	1.14	8.63	9.77	8.46	84.8	
P4940-06	MBHHT2	6	1.16	8.77	9.93	7.87	76.5	
P4940-07	MBHHT3	7	1.15	8.64	9.79	7.98	79.1	
P4940-08	MBHHT4	8	1.12	8.85	9.97	8.34	81.6	
P4940-09	MBHHT5	9	1.16	8.72	9.88	8.57	85.0	
P4940-10	MBHHW1	10	1.16	8.43	9.59	8.41	86.0	
P4940-11	MBHHW2	11	1.18	8.65	9.83	8.75	87.5	
P4940-12	MBHHW3	12	1.18	8.63	9.81	8.82	88.5	
P4940-13	MBHHW4	13	1.19	8.53	9.72	8.8	89.2	
P4940-14	MBHHW5	14	1.18	8.41	9.59	8.73	89.8	
P4940-15	MBHHX7	15	1.16	8.50	9.66	8.52	86.6	
P4940-16	MBHHX8	16	1.15	8.81	9.96	8.11	79.0	
P4940-17	MBHHX8D	17	1.15	8.81	9.96	8.11	79.0	
P4940-18	MBHHX8S	18	1.15	8.81	9.96	8.11	79.0	
P4940-19	MBHHX9	19	1.19	8.50	9.69	8.68	88.1	
P4940-20	MBHHY0	20	1.14	8.75	9.89	8.8	87.5	
P4940-21	MBHHY1	21	1.17	8.58	9.75	8.86	89.6	
P4940-22	MBHHY2	22	1.11	8.76	9.87	8.95	89.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$