

LB183581

7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

46186

A= 0.0000e+000

B= 2.1579e-004

C= -1.9631e-002

Rho= 0.9996241

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	✓/✗
0.0	0.000	0.000	0.000	91	0.000	91					
0.2	0.200	0.224	0.024	1129	0.0 %	1129					12
2.5	2.500	2.537	0.037	11846	0.0 %	11846					1
5.0	5.000	4.828	-0.172	22465	0.0 %	22465					3
7.5	7.500	7.665	0.165	35610	0.0 %	35610					2
10.0	10.000	9.947	-0.053	46186	0.0 %	46186					1

LB133581 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Nov 2024 13:37:48

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	St	91	-	0.0000	7471B	PPB	22 Nov 2024 13:37:58	S	Std
0.2 - 1	50.2	1129	-	0.2000	7471B	PPB	22 Nov 2024 13:40:15	S	Std
2.5 - 1	52.5	11846	-	2.5000	7471B	PPB	22 Nov 2024 13:42:33	S	Std
5.0 - 1	55	22465	-	5.0000	7471B	PPB	22 Nov 2024 13:44:49	S	Std
7.5 - 1	57.5	35610	-	7.5000	7471B	PPB	22 Nov 2024 13:49:53	S	Std
10.0 - 1	57.8	46186	-	10.0000	7471B	PPB	22 Nov 2024 13:55:03	S	Std
ICV98 - 1	ICV98	17021	3.6533	-7471B	PPB	PPB	22 Nov 2024 13:57:56	U	SMPL
ICB98 - 1	ICB98	-159	-0.0539	-7471B	PPB	PPB	22 Nov 2024 14:00:12	U	SMPL
CCV37 - 1	CCV37	22640	4.8658	-7471B	PPB	PPB	22 Nov 2024 14:02:30	U	SMPL
CCB37 - 1	CCB37	-188	-0.0602	-7471B	PPB	PPB	22 Nov 2024 14:04:44	U	SMPL
CRA - 1	CRA	1132	0.2246	-7471B	PPB	PPB	22 Nov 2024 14:07:02	U	SMPL
HighStd - 1	HighStd	46860	10.0922	-7471B	PPB	PPB	22 Nov 2024 14:09:17	U	SMPL
ChkStd - 1	ChkStd	33864	7.2878	-7471B	PPB	PPB	22 Nov 2024 14:11:32	U	SMPL
PB165177BL - 1	PB165177BL	-225	-0.0682	-7471B	PPB	PPB	22 Nov 2024 14:13:58	U	SMPL
PB165177BS - 1	PB165177BS	17483	3.7530	-7471B	PPB	PPB	22 Nov 2024 14:16:21	U	SMPL
P4860-01 - 1	DUP-01	161	0.0151	-7471B	PPB	PPB	22 Nov 2024 14:18:37	U	SMPL
P4860-01DUP - 1	DUP-01DUP	445	0.0764	-7471B	PPB	PPB	22 Nov 2024 14:20:56	U	SMPL
P4860-01MS - 1	DUP-01MS	17084	3.6669	-7471B	PPB	PPB	22 Nov 2024 14:23:13	U	SMPL
P4860-01MSD - 1	DUP-01MSD	17542	3.7657	-7471B	PPB	PPB	22 Nov 2024 14:25:30	U	SMPL
P4860-02 - 1	PH2-BOT-001	1306	0.2622	-7471B	PPB	PPB	22 Nov 2024 14:27:47	U	SMPL
CCV38 - 1	CCV38	22048	4.7381	-7471B	PPB	PPB	22 Nov 2024 14:30:07	U	SMPL
CCB38 - 1	CCB38	-294	-0.0831	-7471B	PPB	PPB	22 Nov 2024 14:32:24	U	SMPL
P4860-03 - 1	PH2-BOT-002	1756	0.3593	-7471B	PPB	PPB	22 Nov 2024 14:34:42	U	SMPL
P4860-04 - 1	PH2-BOT-003	2626	0.5470	-7471B	PPB	PPB	22 Nov 2024 14:36:57	U	SMPL
P4860-05 - 1	PH2-BOT-004	432	0.0736	-7471B	PPB	PPB	22 Nov 2024 14:39:13	U	SMPL
P4860-06 - 1	PH2-BOT-009	334	0.0524	-7471B	PPB	PPB	22 Nov 2024 14:41:30	U	SMPL
P4860-07 - 1	PH2-BOT-008	626	0.1155	-7471B	PPB	PPB	22 Nov 2024 14:43:47	U	SMPL
P4860-08 - 1	PH2-BOT-007	735	0.1390	-7471B	PPB	PPB	22 Nov 2024 14:46:03	U	SMPL
P4860-09 - 1	PH2-BOT-006	665	0.1239	-7471B	PPB	PPB	22 Nov 2024 14:48:19	U	SMPL
P4860-10 - 1	PH2-BOT-005	270	0.0386	-7471B	PPB	PPB	22 Nov 2024 14:50:36	U	SMPL
PB165178BL - 1	PB165178BL	55	-0.0078	-7471B	PPB	PPB	22 Nov 2024 14:52:53	U	SMPL
PB165178BS - 1	PB165178BS	17258	3.7044	-7471B	PPB	PPB	22 Nov 2024 14:55:09	U	SMPL
CCV39 - 1	CCV39	21991	4.7258	-7471B	PPB	PPB	22 Nov 2024 14:57:24	U	SMPL
CCB39 - 1	CCB39	-171	-0.0565	-7471B	PPB	PPB	22 Nov 2024 14:59:43	U	SMPL
P4916-01 - 1	TP-1-WC	491	0.0863	-7471B	PPB	PPB	22 Nov 2024 15:02:01	U	SMPL
P4916-05 - 1	TP-2-WC	1024	0.2013	-7471B	PPB	PPB	22 Nov 2024 15:04:16	U	SMPL
P4916-09 - 1	TP-3-WC	320	0.0494	-7471B	PPB	PPB	22 Nov 2024 15:06:32	U	SMPL
P4923-06 - 1	72-11991	633	0.1170	-7471B	PPB	PPB	22 Nov 2024 15:08:48	U	SMPL
P4924-01 - 1	MH-4	551	0.0993	-7471B	PPB	PPB	22 Nov 2024 15:11:03	U	SMPL
P4925-01 - 1	MH-741	932	0.1815	-7471B	PPB	PPB	22 Nov 2024 15:13:19	U	SMPL
P4925-05 - 1	MH-758	2625	0.5468	-7471B	PPB	PPB	22 Nov 2024 15:15:35	U	SMPL
P4929-01 - 1	ARS520	4045	0.8532	-7471B	PPB	PPB	22 Nov 2024 15:17:52	U	SMPL
P4936-01 - 1	PL-01-11202024	2574	0.5358	-7471B	PPB	PPB	22 Nov 2024 15:20:10	U	SMPL
P4938-01 - 1	MH-732	527	0.0941	-7471B	PPB	PPB	22 Nov 2024 15:22:27	U	SMPL
CCV40 - 1	CCV40	23101	4.9653	-7471B	PPB	PPB	22 Nov 2024 15:24:43	U	SMPL
CCB40 - 1	CCB40	-102	-0.0416	-7471B	PPB	PPB	22 Nov 2024 15:26:58	U	SMPL
P4938-05 - 1	MH-734	3172	0.6648	-7471B	PPB	PPB	22 Nov 2024 15:29:17	U	SMPL
P4938-05DUP - 1	MH-734DUP	3452	0.7253	-7471B	PPB	PPB	22 Nov 2024 15:31:33	U	SMPL
P4938-05MS - 1	MH-734MS	21720	4.6673	-7471B	PPB	PPB	22 Nov 2024 15:33:50	U	SMPL
P4938-05MSD - 1	MH-734MSD	23923	5.1427	-7471B	PPB	PPB	22 Nov 2024 15:36:07	U	SMPL
P4860-01LX5 - 1		-204	-0.0637	-7471B	PPB	PPB	22 Nov 2024 15:38:26	U	SMPL
P4860-01A - 1		17711	3.8022	-7471B	PPB	PPB	22 Nov 2024 15:40:45	U	SMPL
P4938-05LX5 - 1		578	0.1051	-7471B	PPB	PPB	22 Nov 2024 15:43:01	U	SMPL
P4938-05A - 1		22058	4.7402	-7471B	PPB	PPB	22 Nov 2024 15:45:21	U	SMPL
CCV41 - 1	CCV41	22566	4.8498	-7471B	PPB	PPB	22 Nov 2024 15:47:39	U	SMPL
CCB41 - 1	CCB41	-90	-0.0391	-7471B	PPB	PPB	22 Nov 2024 15:49:57	U	SMPL