



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/25/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 15:20
In Date: 11/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 11/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133584

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4964-01	MBHJ01	1	1.17	8.49	9.66	8.14	82.1	
P4964-02	MBHJ01D	2	1.17	8.49	9.66	8.14	82.1	
P4964-03	MBHJ01S	3	1.17	8.49	9.66	8.14	82.1	
P4964-04	MBHJ02	4	1.19	8.52	9.71	9.17	93.7	
P4964-05	MBHJ03	5	1.18	8.49	9.67	9.45	97.4	
P4964-06	MBHJ04	6	1.18	8.56	9.74	9.55	97.8	
P4964-07	MBHJ05	7	1.15	8.53	9.68	9.26	95.1	
P4964-08	MBHJ20	8	1.19	8.52	9.71	7.75	77.0	
P4964-09	MBHJ21	9	1.19	8.70	9.89	8.53	84.4	
P4964-10	MBHJ22	10	1.16	8.53	9.69	8.63	87.6	
P4964-11	MBHJ23	11	1.17	8.60	9.77	8.75	88.1	
P4964-12	MBHJ24	12	1.16	8.41	9.57	8.62	88.7	
P4964-13	MBHJ25	13	1.14	8.77	9.91	9.03	90.0	
P4964-14	MBHJ27	14	1.15	8.40	9.55	6.81	67.4	
P4964-15	MBHJ28	15	1.19	8.52	9.71	7.65	75.8	
P4964-16	MBHJ29	16	1.18	8.40	9.58	8.13	82.7	
P4964-17	MBHJ30	17	1.16	8.59	9.75	7.71	76.3	
P4964-18	MBHJ31	18	1.17	8.60	9.77	7.97	79.1	
P4964-19	MBHJ32	19	1.14	8.61	9.75	8.2	82.0	
P4964-20	MBHJ33	20	1.19	8.37	9.56	8.19	83.6	
P4964-21	MBHJ41	21	1.16	8.61	9.77	8.19	81.6	
P4964-22	MBHJ42	22	1.18	8.51	9.69	8.17	82.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$