

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133589

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/22/2024	15:30
2	CAL2	1	Water	NA	NA	20.3	7.00	11/22/2024	15:31
3	CAL3	1	Water	NA	NA	20.3	10.03	11/22/2024	15:33
4	ICV	1	Water	NA	NA	20.3	7.00	11/22/2024	15:38
5	CCV1	1	Water	NA	NA	20.3	2.01	11/22/2024	15:40
6	P4970-02	1	Water	NA	NA	20.7	7.85	11/22/2024	15:47
7	P4970-02DUP	1	Water	NA	NA	20.8	7.86	11/22/2024	15:48
8	P4972-01	1	Water	NA	NA	20.1	7.54	11/22/2024	15:50
9	P4972-02	1	Water	NA	NA	20.4	7.71	11/22/2024	16:00
10	P4972-03	1	Water	NA	NA	20.5	7.66	11/22/2024	16:05
11	P4972-05	1	Water	NA	NA	20.7	5.34	11/22/2024	16:15
12	P4974-01	1	Water	NA	NA	20.6	6.37	11/22/2024	16:20
13	CCV2	1	Water	NA	NA	20.3	12.02	11/22/2024	16:24

WORKLIST(Hardcopy Internal Chain)

VB 133589

WorkList Name : PH WATER P4974 WorkList ID : 185713 Department : Wet-Chemistry Date : 11-22-2024 15:44:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4970-02	60235	Water	pH	Cool 4 deg C	PSEG03	L11	11/22/2024	9040C
P4972-01	111524-A	Water	pH	Cool 4 deg C	PSEG03	L41	11/22/2024	9040C
P4972-02	111524-B	Water	pH	Cool 4 deg C	PSEG03	L41	11/22/2024	9040C
P4972-03	912-A	Water	pH	Cool 4 deg C	PSEG03	L41	11/22/2024	9040C
P4972-05	628-F	Water	pH	Cool 4 deg C	PSEG03	L41	11/22/2024	9040C
P4974-01	111824-DL	Water	pH	Cool 4 deg C	PSEG03	L61	11/22/2024	9040C

Date/Time 11-22-24 15:00
 Raw Sample Received by: ZB WGC
 Raw Sample Relinquished by: ZB WGC

Date/Time 11-22-24 17:30
 Raw Sample Received by: ZB WGC
 Raw Sample Relinquished by: ZB WGC