

LB133592


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4020E



A= 0.0000e+000

B= 2.4890e-004 slope

C= -6.2742e-002 y-intercept

Rho= 0.9999183

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	-0.016	-0.016	188	0.000	188					
0.05	0.050					442	0				
0.20	0.200	0.207	0.007	1084	0.0 %	1084					
2.50	2.500	2.506	0.006	10320	0.0 %	10320					
5.00	5.000	4.965	-0.035	20201	0.0 %	20201					
7.50	7.500	7.593	0.093	30760	0.0 %	30760					
10.0	10.000	9.944	-0.056	40206	0.0 %	40206					

LB133592

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	188	-		0 SFAM01.1	PPB	11/22/2024 16:45	Std
	0.2 S01	1084	-		0.2 SFAM01.1	PPB	11/22/2024 16:47	Std
	2.5 S02	10320	-		2.5 SFAM01.1	PPB	11/22/2024 16:50	Std
	5 S03	20201	-		5 SFAM01.1	PPB	11/22/2024 16:52	Std
	7.5 S04	30760	-		7.5 SFAM01.1	PPB	11/22/2024 16:54	Std
	10 S05	40206	-		10 SFAM01.1	PPB	11/22/2024 17:02	Std
ICV099	ICV099	16377	4.0135 -		SFAM01.1	PPB	11/22/2024 17:05	SMPL
ICB099	ICB099	28	-0.0558 -		SFAM01.1	PPB	11/22/2024 17:07	SMPL
CCV042	CCV042	21245	5.2251 -		SFAM01.1	PPB	11/22/2024 17:09	SMPL
CCB042	CCB042	31	-0.055 -		SFAM01.1	PPB	11/22/2024 17:12	SMPL
PB165207BL	PBS207	115	-0.0341 -		SFAM01.1	PPB	11/22/2024 17:14	SMPL
P4728-01	MBH9B8	3569	0.8256 -		SFAM01.1	PPB	11/22/2024 17:16	SMPL
P4728-02	MBH9B8D	3242	0.7442 -		SFAM01.1	PPB	11/22/2024 17:19	SMPL
P4728-03	MBH9B8S	14494	3.5448 -		SFAM01.1	PPB	11/22/2024 17:21	SMPL
P4728-04	MBH9B9	694	0.11 -		SFAM01.1	PPB	11/22/2024 17:23	SMPL
P4728-05	MBH9C0	172	-0.0199 -		SFAM01.1	PPB	11/22/2024 17:25	SMPL
P4728-06	MBH9C2	1641	0.3457 -		SFAM01.1	PPB	11/22/2024 17:28	SMPL
P4728-07	MBH9C3	715	0.1152 -		SFAM01.1	PPB	11/22/2024 17:30	SMPL
P4728-08	MBH9C4	551	0.0744 -		SFAM01.1	PPB	11/22/2024 17:32	SMPL
P4728-09	MBH9D4	1078	0.2056 -		SFAM01.1	PPB	11/22/2024 17:34	SMPL
P4728-10	MBH9D5	509	0.0639 -		SFAM01.1	PPB	11/22/2024 17:37	SMPL
P4728-11	MBH9D6	315	0.0157 -		SFAM01.1	PPB	11/22/2024 17:39	SMPL
P4728-12	MBH9F0	2446	0.5461 -		SFAM01.1	PPB	11/22/2024 17:41	SMPL
P4728-13	MBH9F1	912	0.1643 -		SFAM01.1	PPB	11/22/2024 17:43	SMPL
P4728-14	MBH9F2	334	0.0204 -		SFAM01.1	PPB	11/22/2024 17:46	SMPL
P4728-15	MBH9G6	2522	0.565 -		SFAM01.1	PPB	11/22/2024 17:48	SMPL
P4728-17	MBH9F4	2764	0.6252 -		SFAM01.1	PPB	11/22/2024 17:50	SMPL
P4728-18	MBH9F5	586	0.0831 -		SFAM01.1	PPB	11/22/2024 17:52	SMPL
P4728-19	MBH9F6	423	0.0425 -		SFAM01.1	PPB	11/22/2024 17:55	SMPL
P4728-20	MBH9F8	2800	0.6342 -		SFAM01.1	PPB	11/22/2024 17:57	SMPL
CCV043	CCV043	21679	5.3331 -		SFAM01.1	PPB	11/22/2024 17:59	SMPL
CCB043	CCB043	74	-0.0443 -		SFAM01.1	PPB	11/22/2024 18:02	SMPL
P4728-21	MBH9F9	382	0.0323 -		SFAM01.1	PPB	11/22/2024 18:04	SMPL
P4728-22	MBH9G0	325	0.0181 -		SFAM01.1	PPB	11/22/2024 18:06	SMPL
PB165209BL	PBS209	222	-0.0075 -		SFAM01.1	PPB	11/22/2024 18:08	SMPL
P4750-01	MBH9C6	2154	0.4734 -		SFAM01.1	PPB	11/22/2024 18:11	SMPL
P4750-02	MBH9C7	329	0.0191 -		SFAM01.1	PPB	11/22/2024 18:13	SMPL
P4750-03	MBH9C8	296	0.0109 -		SFAM01.1	PPB	11/22/2024 18:15	SMPL
P4750-04	MBH9D0	1739	0.3701 -		SFAM01.1	PPB	11/22/2024 18:17	SMPL
P4750-05	MBH9D1	358	0.0264 -		SFAM01.1	PPB	11/22/2024 18:20	SMPL
P4750-06	MBH9D2	319	0.0167 -		SFAM01.1	PPB	11/22/2024 18:22	SMPL
P4750-07	MBH9D8	2199	0.4846 -		SFAM01.1	PPB	11/22/2024 18:24	SMPL
P4750-08	MBH9D9	547	0.0734 -		SFAM01.1	PPB	11/22/2024 18:27	SMPL
P4750-09	MBH9E0	390	0.0343 -		SFAM01.1	PPB	11/22/2024 18:29	SMPL
P4750-10	MBH9E2	1208	0.2379 -		SFAM01.1	PPB	11/22/2024 18:31	SMPL
P4750-11	MBH9E3	270	0.0045 -		SFAM01.1	PPB	11/22/2024 18:33	SMPL

LB133592

INSTRUMENT ID : CV1

P4750-12	MBH9E4	123	-0.0321 -	SFAM01.1	PPB	11/22/2024 18:36 SMPL
P4750-13	MBH9E6	3077	0.7031 -	SFAM01.1	PPB	11/22/2024 18:38 SMPL
P4750-14	MBH9E6D	3172	0.7268 -	SFAM01.1	PPB	11/22/2024 18:40 SMPL
P4750-15	MBH9E6S	13183	3.2185 -	SFAM01.1	PPB	11/22/2024 18:43 SMPL
P4750-16	MBH9E7	175	-0.0192 -	SFAM01.1	PPB	11/22/2024 18:45 SMPL
P4750-17	MBH9E8	332	0.0199 -	SFAM01.1	PPB	11/22/2024 18:47 SMPL
CCV044	CCV044	21823	5.3689 -	SFAM01.1	PPB	11/22/2024 18:49 SMPL
CCB044	CCB044	84	-0.0418 -	SFAM01.1	PPB	11/22/2024 18:52 SMPL
P4750-18	MBH9G2	1742	0.3708 -	SFAM01.1	PPB	11/22/2024 18:54 SMPL
P4750-19	MBH9G3	934	0.1697 -	SFAM01.1	PPB	11/22/2024 18:56 SMPL
P4750-20	MBH9G4	363	0.0276 -	SFAM01.1	PPB	11/22/2024 18:59 SMPL
P4750-21	MBH9G7	332	0.0199 -	SFAM01.1	PPB	11/22/2024 19:01 SMPL
CCV045	CCV045	22060	5.4279 -	SFAM01.1	PPB	11/22/2024 19:06 SMPL
CCB045	CCB045	58	-0.0483 -	SFAM01.1	PPB	11/22/2024 19:09 SMPL