

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133594

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	11/22/2024	14:10
2	P4947-01		1	11	100.00	>212.0	>212.0	11/22/2024	14:40
3	P4947-01DUP		1	11	100.00	>212.0	>212.0	11/22/2024	15:10
4	P4949-01		1	14	100.00	>212.0	>212.0	11/22/2024	15:40
5	P4950-01		1	12	100.00	>212.0	>212.0	11/22/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-11-22

WorkList ID : 185694

Department : Wet-Chemistry

Date : 11-22-2024 11:39:37

LB133594

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4947-01	A3988	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4949-01	REACTOR-SUMP	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4950-01	AUD-1356	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B

Date/Time 11/22/2024 13.00
Raw Sample Received by: RM cws
Raw Sample Relinquished by: JH Wey

Date/Time 11/22/2024 16.50
Raw Sample Received by: JH Wey
Raw Sample Relinquished by: RM cws