

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133598

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	11/25/2024	08:20
2	P4952-04		1	11	100.00	>212.0	>212.0	11/25/2024	08:50
3	P4952-04DUP		1	11	100.00	>212.0	>212.0	11/25/2024	09:20
4	P4952-07		1	5	8.00	46.4	45.7	11/25/2024	09:50
5	P4970-01		1	5	7.00	44.6	43.9	11/25/2024	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133598

WorkList Name : fp-s-11-25

WorkList ID : 185719

Department : Wet-Chemistry

Date : 11-25-2024 08:03:01

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4952-04 A	GAS-BUR-1614	Solid	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4952-07 A	GAS-BUR-1621	Solid	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4970-01 A	34565	Solid	Flash Point	Cool 4 deg C	PSEG03	L11	11/22/2024	1010B

Date/Time 11/25/2024 08:10
Raw Sample Received by: RY CWG
Raw Sample Relinquished by: JP WPC

Date/Time 11/25/2024 11:00
Raw Sample Received by: JP WPC
Raw Sample Relinquished by: RY CWG