



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 11/26/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:30
In Date: 11/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 11/26/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133609

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4980-01	MBHJ70	1	1.15	6.93	8.08	7.14	86.4	
P4980-02	MBHJ71	2	1.18	8.51	9.69	9.19	94.1	
P4980-03	MBHJ72	3	1.14	8.63	9.77	9.28	94.3	
P4980-04	MBHJ73	4	1.14	8.82	9.96	8.56	84.1	
P4980-05	MBHJ74	5	1.11	8.78	9.89	8.43	83.4	
P4980-06	MBHJ75	6	1.15	8.82	9.97	8.38	82.0	
P4980-07	MBHJ76	7	1.13	8.81	9.94	8.46	83.2	
P4980-08	MBHJ84	8	1.16	8.51	9.67	7.44	73.8	
P4980-09	MBHJ85	9	1.13	8.71	9.84	8.1	80.0	
P4980-10	MBHJ86	10	1.17	8.80	9.97	9.47	94.3	
P4980-11	MBHJ87	11	1.16	8.81	9.97	8.85	87.3	
P4980-12	MBHJ88	12	1.17	8.55	9.72	8.32	83.6	
P4980-13	MBHJ88D	13	1.17	8.55	9.72	8.32	83.6	
P4980-14	MBHJ88S	14	1.17	8.55	9.72	8.32	83.6	
P4980-15	MBHJ89	15	1.16	8.57	9.73	8.35	83.9	
P4980-16	MBHJ90	16	1.16	8.63	9.79	9.00	90.8	
P4980-17	MBHJB0	17	1.16	8.40	9.56	9.04	93.8	
P4980-18	MBHJB1	18	1.1	8.79	9.89	9.2	92.2	
P4980-19	MBHJB2	19	1.17	8.60	9.77	8.52	85.5	
P4980-20	MBHJB3	20	1.15	8.70	9.85	8.64	86.1	
P4980-21	MBHJB4	21	1.14	8.70	9.84	8.36	83.0	
P4980-22	MBHJB5	22	1.16	8.78	9.94	8.17	79.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$