


LB133612

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 2.5893e-004
 C= -8.8986e-002
 Rho= 0.9990780
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/d
0.0	0.000	-0.090	-0.090	-4	0.000	-4					-10
0.2	0.200	0.180	-0.020	1039	0.0 %	1039					4
2.5	2.500	2.596	0.096	10371	0.0 %	10371					-1
5.0	5.000	4.965	-0.035	19518	0.0 %	19518					5
7.5	7.500	7.780	0.280	30389	0.0 %	30389					-2
10.0	10.000	9.769	-0.231	38072	0.0 %	38072					

LB133612 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Nov 2024 14:02:12

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-4	-	0.000	7471B	PPB	25 Nov 2024 14:13:03	S	Std
0.2 - 1	50.2	1039	-	0.200	7471B	PPB	25 Nov 2024 14:15:19	S	Std
2.5 - 1	52.5	10371	-	2.500	7471B	PPB	25 Nov 2024 14:17:37	S	Std
5.0 - 1	55	19518	-	5.000	7471B	PPB	25 Nov 2024 14:22:41	S	Std
7.5 - 1	57.5	30389	-	7.500	7471B	PPB	25 Nov 2024 14:24:58	S	Std
10.0 - 1	50	38072	-	10.000	7471B	PPB	25 Nov 2024 14:30:02	S	Std
ICV101 - 1	ICV101	14614	3.6950	-7471B	PPB	25 Nov 2024 14:32:53	U	SMPL	
ICB101 - 1	ICB101	94	-0.0646	-7471B	PPB	25 Nov 2024 14:38:07	U	SMPL	
CCV49 - 1	CCV49	19857	5.0526	-7471B	PPB	25 Nov 2024 14:40:21	U	SMPL	
CCB49 - 1	CCB49	124	-0.0569	-7471B	PPB	25 Nov 2024 14:45:22	U	SMPL	
CRA - 1	CRA	1004	0.1710	-7471B	PPB	25 Nov 2024 14:47:37	U	SMPL	
HighStd - 1	HighStd	38474	9.8731	-7471B	PPB	25 Nov 2024 14:49:52	U	SMPL	
ChkStd - 1	ChkStd	26138	6.6790	-7471B	PPB	25 Nov 2024 14:52:07	U	SMPL	
PB165236BL - 1	PBS	-8	-0.0911	-7471B	PPB	25 Nov 2024 14:54:31	U	SMPL	
PB165236BS - 1	LCSS	16358	4.1466	-7471B	PPB	25 Nov 2024 14:56:51	U	SMPL	
P4948-01 - 1	337	2690	0.6075	-7471B	PPB	25 Nov 2024 14:59:07	U	SMPL	
P4948-03 - 1	72-11944	2901	0.6622	-7471B	PPB	25 Nov 2024 15:01:26	U	SMPL	
P4949-03 - 1	CONTAINMENT-STONE	220	-0.0320	-7471B	PPB	25 Nov 2024 15:03:46	U	SMPL	
P4951-01 - 1	AU-05-112124	1356	0.2621	-7471B	PPB	25 Nov 2024 15:06:05	U	SMPL	
P4954-01 - 1	TR-05-112124	1553	0.3131	-7471B	PPB	25 Nov 2024 15:08:20	U	SMPL	
CCV50 - 1	CCV50	19567	4.9775	-7471B	PPB	25 Nov 2024 15:10:36	U	SMPL	
CCB50 - 1	CCB50	238	-0.0274	-7471B	PPB	25 Nov 2024 15:15:28	U	SMPL	
P4954-03 - 1	TR-06-112124	195	-0.0385	-7471B	PPB	25 Nov 2024 15:17:43	U	SMPL	
P4971-01 - 1	CL-02-112224	1872	0.3957	-7471B	PPB	25 Nov 2024 15:25:48	U	SMPL	
P4971-01DUP - 1	CL-02-112224DUP	1747	0.3634	-7471B	PPB	25 Nov 2024 15:28:03	U	SMPL	
P4971-01MS - 1	CL-02-112224MS	17825	4.5265	-7471B	PPB	25 Nov 2024 15:30:18	U	SMPL	
P4971-01MSD - 1	CL-02-112224MSD	18474	4.6945	-7471B	PPB	25 Nov 2024 15:32:34	U	SMPL	
P4971-01LX5 - 1		42	-0.0781	-7471B	PPB	25 Nov 2024 15:34:53	U	SMPL	
P4971-01A - 1		17943	4.5570	-7471B	PPB	25 Nov 2024 15:37:13	U	SMPL	
CCV51 - 1	CCV51	19461	4.9501	-7471B	PPB	25 Nov 2024 15:39:30	U	SMPL	
CCB51 - 1	CCB51	148	-0.0507	-7471B	PPB	25 Nov 2024 15:41:50	U	SMPL	