

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133617

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	11/25/2024	11:15
2	P4952-02		1	13	100.00	>212.0	>212.0	11/25/2024	11:45
3	P4952-02DUP		1	13	100.00	>212.0	>212.0	11/25/2024	12:15
4	P4952-03		1	14	100.00	>212.0	>212.0	11/25/2024	12:45
5	P4953-01		1	11	100.00	>212.0	>212.0	11/25/2024	14:05
6	P4970-02		1	14	100.00	>212.0	>212.0	11/25/2024	14:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6133617

WorkList Name : fp 1125

WorkList ID : 185716

Department : Wet-Chemistry

Date : 11-25-2024 07:58:45

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4952-02	GAS-BUR-1617	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4952-03	GAS-BUR-1618	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4953-01	1297	Water	Flash Point	Cool 4 deg C	PSEG03	L61	11/21/2024	1010B
P4970-02	60235	Water	Flash Point	Cool 4 deg C	PSEG03	L11	11/22/2024	1010B

Date/Time 11/25/2024 11:25
Raw Sample Received by: RM (Ved)
Raw Sample Relinquished by: JAL (Ved)

Date/Time 11/25/2024 15:05
Raw Sample Received by: JH CEC
Raw Sample Relinquished by: RM CEC