


LB 133631

7470A

INSTRUMENT ID: C12



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.023	0.023	222	0.000	222	0				0.000
0.2	0.200	0.203	0.003	934	0.0 %	934					2
2.5	2.500	2.434	-0.066	9808	0.0 %	9808					-3
5.0	5.000	4.959	-0.041	19849	0.0 %	19849					1
7.5	7.500	7.700	0.200	30746	0.0 %	30746					3
10.0	10.000	9.881	-0.119	39421	0.0 %	39421					-1

LB133631 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Nov 2024 11:06:12

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	56	222	-	0.0000	7470A	PPB	26 Nov 2024 11:18:08	S	Std
0.2 - 1	56.2	934	-	0.2000	7470A	PPB	26 Nov 2024 11:20:24	S	Std
2.5 - 1	52.5	9808	-	2.5000	7470A	PPB	26 Nov 2024 11:22:41	S	Std
5.0 - 1	55	19849	-	5.0000	7470A	PPB	26 Nov 2024 11:24:58	S	Std
7.5 - 1	52.5	30746	-	7.5000	7470A	PPB	26 Nov 2024 11:27:15	S	Std
10.0 - 1	510	39421	-	10.0000	7470A	PPB	26 Nov 2024 11:29:32	S	Std
ICV102 - 1	ICV102	15338	3.8248	-	7470A	PPB	26 Nov 2024 11:32:30	U	SMPL
ICB102 - 1	ICB102	-120	-0.0625	-	7470A	PPB	26 Nov 2024 11:34:44	U	SMPL
CCV52 - 1	CCV52	20285	5.0689	-	7470A	PPB	26 Nov 2024 11:37:01	U	SMPL
CCB52 - 1	CCB52	-157	-0.0718	-	7470A	PPB	26 Nov 2024 11:39:15	U	SMPL
CRA - 1	CRA	833	0.1771	-	7470A	PPB	26 Nov 2024 11:41:32	U	SMPL
HighStd - 1	HighStd	38964	9.7662	-	7470A	PPB	26 Nov 2024 11:43:47	U	SMPL
ChkStd - 1	ChkStd	29229	7.3181	-	7470A	PPB	26 Nov 2024 11:46:02	U	SMPL
PB165265BL - 1	PBW	-116	-0.0615	-	7470A	PPB	26 Nov 2024 11:52:48	U	SMPL
PB165265BS - 1	LCSW	18038	4.5038	-	7470A	PPB	26 Nov 2024 11:55:03	U	SMPL
P4938-04 - 1	MH-732	-197	-0.0819	-	7470A	PPB	26 Nov 2024 11:57:20	U	SMPL
P4938-04DUP - 1	MH-732DUP	-35	-0.0411	-	7470A	PPB	26 Nov 2024 11:59:37	U	SMPL
P4938-04MS - 1	MH-732MS	16824	4.1985	-	7470A	PPB	26 Nov 2024 12:01:54	U	SMPL
P4938-04MSD - 1	MH-732MSD	16002	3.9918	-	7470A	PPB	26 Nov 2024 12:04:10	U	SMPL
P4938-08 - 1	MH-734	-308	-0.1098	-	7470A	PPB	26 Nov 2024 12:06:27	U	SMPL
CCV53 - 1	CCV53	19284	4.8171	-	7470A	PPB	26 Nov 2024 12:08:45	U	SMPL
CCB53 - 1	CCB53	-196	-0.0816	-	7470A	PPB	26 Nov 2024 12:11:00	U	SMPL
P4946-01 - 1	BW-1	-16	-0.0364	-	7470A	PPB	26 Nov 2024 12:13:17	U	SMPL
P4946-02 - 1	BW-2	3336	0.8066	-	7470A	PPB	26 Nov 2024 12:15:32	U	SMPL
PB165159TB - 1	PB165159TB	-128	-0.0645	-	7470A	PPB	26 Nov 2024 12:17:47	U	SMPL
P4938-04LX5 - 1		-154	-0.0711	-	7470A	PPB	26 Nov 2024 12:20:05	U	SMPL
P4938-04A - 1		16423	4.0976	-	7470A	PPB	26 Nov 2024 12:22:21	U	SMPL
CCV54 - 1	CCV54	19448	4.8584	-	7470A	PPB	26 Nov 2024 12:27:18	U	SMPL
CCB54 - 1	CCB54	-29	-0.0396	-	7470A	PPB	26 Nov 2024 12:29:35	U	SMPL