

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB133641

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	11/26/2024	10:15
2	P4973-06		1	11	100.00	>212.0	>212.0	11/26/2024	10:45
3	P4973-06DUP		1	11	100.00	>212.0	>212.0	11/26/2024	11:15
4	P4974-01		1	13	100.00	>212.0	>212.0	11/26/2024	13:05
5	P4993-01		1	11	100.00	>212.0	>212.0	11/26/2024	13:35
6	P5013-02		1	14	100.00	>212.0	>212.0	11/26/2024	14:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06