

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB133647

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/26/2024	15:35
2	CAL2	1	Water	NA	NA	20.3	7.00	11/26/2024	15:36
3	CAL3	1	Water	NA	NA	20.3	10.02	11/26/2024	15:38
4	ICV	1	Water	NA	NA	20.3	7.01	11/26/2024	15:40
5	CCV1	1	Water	NA	NA	20.3	2.01	11/26/2024	15:44
6	P5006-18	1	Water	NA	NA	20.7	6.10	11/26/2024	15:57
7	P5006-18DUP	1	Water	NA	NA	20.8	6.11	11/26/2024	15:58
8	P5013-02	1	Water	NA	NA	20.7	6.67	11/26/2024	16:10
9	CCV2	1	Water	NA	NA	20.3	12.02	11/26/2024	16:15

WORKLIST(Hardcopy Internal Chain)

VB133647

WorkList Name : ph p5006

WorkList ID : 185806

Department : Wet-Chemistry

Date : 11-26-2024 15:24:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5006-18 B	TW-WTS-02	Water	pH	Cool 4 deg C	ENTA05	L51	11/25/2024	9040C
P5013-02 A	SPILL-1125-B	Water	pH	Cool 4 deg C	PSEG03	L51	11/26/2024	9040C

Date/Time 11-26-24 15:30
Raw Sample Received by: TD (SM)
Raw Sample Relinquished by: TD (SM)

Date/Time 11-26-24 17:30
Raw Sample Received by: TD (SM)
Raw Sample Relinquished by: TD (SM)