

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133652

Slope : 98.3

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/27/2024	08:05
2	CAL2	1	Water	NA	NA	20.2	7.00	11/27/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.02	11/27/2024	08:11
4	ICV	1	Water	NA	NA	20.3	7.01	11/27/2024	08:15
5	CCV1	1	Water	NA	NA	20.3	2.01	11/27/2024	08:17
6	P5013-01	1	Solid	20.02	20	24.6	6.41	11/27/2024	08:25
7	P5013-01DUP	1	Solid	20.03	20	24.7	6.42	11/27/2024	08:26
8	CCV2	1	Water	NA	NA	20.3	12.02	11/27/2024	08:30

WORKLIST(Hardcopy Internal Chain)

UP 133652

WorkList Name : ph s p5013

WorkList ID : 185819

Department : Wet-Chemistry

Date : 11-27-2024 07:47:27

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5013-01	SPILL-1125-A	Solid	pH	Cool 4 deg C	PSEG03	L51	11/26/2024	9045D

Date/Time 11-27-24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11-27-24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]