


LB 133657

7471B

INSTRUMENT ID: C109

 μ Abs.:

45262



Linear

A= 0.0000e+000

B= 2.1952e-004

C= -4.8275e-002

Rho= 0.9997522

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.019	-0.019	133	0.000	133					1
0.2	0.200	0.205	0.005	1152	0.0 %	1152					2
2.5	2.500	2.484	-0.016	11537	0.0 %	11537					1
5.0	5.000	4.979	-0.021	22901	0.0 %	22901					0
7.5	7.500	7.664	0.164	35131	0.0 %	35131					2
10.0	10.000	9.888	-0.112	45262	0.0 %	45262					1

LB133657 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Nov 2024 11:38:46

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	133	-	0.0000	7471B	PPB	27 Nov 2024 12:04:54	S	Std
0.2 - 1	50.2	1152	-	0.2000	7471B	PPB	27 Nov 2024 12:07:11	S	Std
2.5 - 1	52.5	11537	-	2.5000	7471B	PPB	27 Nov 2024 12:09:28	S	Std
5.0 - 1	55	22901	-	5.0000	7471B	PPB	27 Nov 2024 12:11:45	S	Std
7.5 - 1	57.5	35131	-	7.5000	7471B	PPB	27 Nov 2024 12:22:07	S	Std
10.0 - 1	510	45262	-	10.0000	7471B	PPB	27 Nov 2024 12:24:24	S	Std
ICV104 - 1	ICV104	19122	4.1494	-	7471B	PPB	27 Nov 2024 12:27:26	U	SMPL
ICB104 - 1	ICB104	-177	-0.0871	-	7471B	PPB	27 Nov 2024 12:29:41	U	SMPL
CCV59 - 1	CCV59	22443	4.8784	-	7471B	PPB	27 Nov 2024 12:31:59	U	SMPL
CCB59 - 1	CCB59	48	-0.0377	-	7471B	PPB	27 Nov 2024 12:36:53	U	SMPL
CRA - 1	CRA	1086	0.1901	-	7471B	PPB	27 Nov 2024 12:39:08	U	SMPL
HighStd - 1	HighStd	45481	9.9357	-	7471B	PPB	27 Nov 2024 12:41:22	U	SMPL
ChkStd - 1	ChkStd	32181	7.0161	-	7471B	PPB	27 Nov 2024 12:43:37	U	SMPL
PB165303BL - 1	PB165303BL	68	-0.0333	-	7471B	PPB	27 Nov 2024 12:48:48	U	SMPL
PB165303BS - 1	PB165303BS	19341	4.1974	-	7471B	PPB	27 Nov 2024 12:51:03	U	SMPL
P4985-01 - 1	MH-756-WC	633	0.0907	-	7471B	PPB	27 Nov 2024 12:53:19	U	SMPL
P4985-05 - 1	MH-740-WC	2029	0.3971	-	7471B	PPB	27 Nov 2024 12:56:37	U	SMPL
P4986-01 - 1	TAPIAL1-SB04D-9-112224-00-T1	487	0.0586	-	7471B	PPB	27 Nov 2024 13:00:15	U	SMPL
P4986-01DUP - 1	TAPIAL1-SB04D-9-112224-00-T1DUP	872	0.1431	-	7471B	PPB	27 Nov 2024 13:07:04	U	SMPL
P4986-01MS - 1	TAPIAL1-SB04D-9-112224-00-T1MS	20130	4.3706	-	7471B	PPB	27 Nov 2024 13:09:22	U	SMPL
CCV60 - 1	CCV60	22617	4.9166	-	7471B	PPB	27 Nov 2024 13:11:37	U	SMPL
CCB60 - 1	CCB60	-226	-0.0979	-	7471B	PPB	27 Nov 2024 13:13:59	U	SMPL
P4986-01MSD - 1	TAPIAL1-SB04D-9-112224-00-T1MSD	19680	4.2675	-	7471B	PPB	27 Nov 2024 13:16:15	U	SMPL
P4986-02 - 1	TAPIAL1-SB04D-4-112224-00-T1	510	0.0637	-	7471B	PPB	27 Nov 2024 13:18:31	U	SMPL
P5000-01 - 1	MH-745	511	0.0639	-	7471B	PPB	27 Nov 2024 13:20:49	U	SMPL
P5000-05 - 1	MH-733	11241	2.4193	-	7471B	PPB	27 Nov 2024 13:23:06	U	SMPL
P5005-01 - 1	STOCK-PILE	432	0.0466	-	7471B	PPB	27 Nov 2024 13:25:24	U	SMPL
P5019-01 - 1	EO-02-11262024	5294	1.1139	-	7471B	PPB	27 Nov 2024 13:27:41	U	SMPL
P4986-01LX5 - 1		-101	-0.0704	-	7471B	PPB	27 Nov 2024 13:29:58	U	SMPL
P4986-01A - 1		19842	4.3074	-	7471B	PPB			
CCV61 - 1	CCV61	23198	5.0441	-	7471B	PPB			
CCB61 - 1	CCB61	-163	-0.0841	-	7471B	PPB			