



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/2/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 11/27/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:44
Out Date: 11/28/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133659

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5011-01	MBHJM6	1	1.15	8.52	9.67	7.58	75.5	
P5011-02	MBHJM7	2	1.17	8.60	9.77	8.22	82.0	
P5011-03	MBHJM8	3	1.15	8.63	9.78	8.25	82.3	
P5011-04	MBHJM9	4	1.16	8.78	9.94	8.28	81.1	
P5011-05	MBHJN0	5	1.18	8.48	9.66	8.03	80.8	
P5011-06	MBHJN1	6	1.19	8.52	9.71	8.07	80.8	
P5011-07	MBHJN2	7	1.15	8.40	9.55	8.13	83.1	
P5011-08	MBHJN3	8	1.14	8.40	9.54	7.48	75.5	
P5011-09	MBHJN4	9	1.16	8.63	9.79	8.11	80.5	
P5011-10	MBHJN5	10	1.18	8.62	9.8	8.55	85.5	
P5011-11	MBHJN6	11	1.19	8.42	9.61	8.35	85.0	
P5011-12	MBHJN7	12	1.16	8.44	9.6	8.17	83.1	
P5011-13	MBHJN8	13	1.14	8.78	9.92	8.45	83.3	
P5011-14	MBHJN9	14	1.16	8.49	9.65	8.66	88.3	
P5011-15	MBHJP0	15	1.15	8.77	9.92	8.13	79.6	
P5011-16	MBHJP1	16	1.12	8.77	9.89	8.63	85.6	
P5011-17	MBHJP2	17	1.16	8.50	9.66	8.49	86.2	
P5011-18	MBHJP3	18	1.15	8.40	9.55	8.33	85.5	
P5011-19	MBHJP4	19	1.18	8.53	9.71	8.55	86.4	
P5011-20	MBHJP5	20	1.18	8.66	9.84	8.84	88.5	
P5011-21	MBHJP5D	21	1.18	8.66	9.84	8.84	88.5	
P5011-22	MBHJP5S	22	1.18	8.66	9.84	8.84	88.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$