

LA133664


7470A

INSTRUMENT ID: CX1

Linear

μ Abs.:

4501E



A= 0.0000e+000

B= 2.1709e-004

C= -1.1419e-001

Rho= 0.9982739

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.078	-0.078	168	0.000	168				
0.2	0.200	0.188	-0.012	1394	0.0 %	1394				
2.5	2.500	2.504	0.004	12062	0.0 %	12062				
5.0	5.000	5.029	0.029	23693	0.0 %	23693				
7.5	7.500	7.897	0.397	36905	0.0 %	36905				
10.0	10.000	9.658	-0.342	45016	0.0 %	45016				

6/21
-6
0
1
5
3

LB133664 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Nov 2024 15:22:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	168	-	0.0000	7470A	PPB	27 Nov 2024 15:29:34	S	Std
0.2 - 1	50.2	1394	-	0.2000	7470A	PPB	27 Nov 2024 15:31:50	S	Std
2.5 - 1	52.5	12062	-	2.5000	7470A	PPB	27 Nov 2024 15:34:07	S	Std
5.0 - 1	55	23693	-	5.0000	7470A	PPB	27 Nov 2024 15:36:24	S	Std
7.5 - 1	59.5	36905	-	7.5000	7470A	PPB	27 Nov 2024 15:41:20	S	Std
10.0 - 1	510	45016	-	10.0000	7470A	PPB	27 Nov 2024 15:53:09	S	Std
ICV105 - 1	ICV105	19588	4.1381	-	7470A	PPB	27 Nov 2024 15:59:40	U	SMPL
ICB105 - 1	ICB105	281	-0.0532	-	7470A	PPB	27 Nov 2024 16:01:55	U	SMPL
CCV62 - 1	CCV62	23845	5.0623	-	7470A	PPB	27 Nov 2024 16:04:14	U	SMPL
CCB62 - 1	CCB62	210	-0.0686	-	7470A	PPB	27 Nov 2024 16:06:29	U	SMPL
CRA - 1	CRA	1412	0.1923	-	7470A	PPB	27 Nov 2024 16:11:22	U	SMPL
HighStd - 1	HighStd	45500	9.7633	-	7470A	PPB	27 Nov 2024 16:13:37	U	SMPL
ChkStd - 1	ChkStd	36825	7.8801	-	7470A	PPB	27 Nov 2024 16:15:52	U	SMPL
PB165267BL - 1	PBW	133	-0.0853	-	7470A	PPB	27 Nov 2024 16:18:10	U	SMPL
PB165267BS - 1	LCSW	18308	3.8603	-	7470A	PPB	27 Nov 2024 16:23:04	U	SMPL
P4985-04 - 1	MH-756-WC	212	-0.0682	-	7470A	PPB	27 Nov 2024 16:25:19	U	SMPL
P4985-08 - 1	MH-740-WC	208	-0.0690	-	7470A	PPB	27 Nov 2024 16:27:37	U	SMPL
P5000-04 - 1	MH-745	172	-0.0769	-	7470A	PPB	27 Nov 2024 16:29:54	U	SMPL
P5000-08 - 1	MH-733	221	-0.0662	-	7470A	PPB	27 Nov 2024 16:32:11	U	SMPL
P5000-08DUP - 1	MH-733DUP	976	0.0977	-	7470A	PPB	27 Nov 2024 16:34:27	U	SMPL
CCV63 - 1	CCV63	23977	5.0909	-	7470A	PPB	27 Nov 2024 16:36:42	U	SMPL
CCB63 - 1	CCB63	340	-0.0404	-	7470A	PPB	27 Nov 2024 16:38:57	U	SMPL
P5000-08MS - 1	MH-733MS	16104	3.3818	-	7470A	PPB	27 Nov 2024 16:41:14	U	SMPL
P5000-08MSD - 1	MH-733MSD	15856	3.3280	-	7470A	PPB	27 Nov 2024 16:43:29	U	SMPL
PB165268BL - 1	PBW	367	-0.0345	-	7470A	PPB	27 Nov 2024 16:45:46	U	SMPL
PB165268BS - 1	LCSW	15939	3.3460	-	7470A	PPB	27 Nov 2024 16:48:04	U	SMPL
P4995-02 - 1	001	710	0.0399	-	7470A	PPB	27 Nov 2024 16:50:19	U	SMPL
P4995-02DUP - 1	001DUP	1030	0.1094	-	7470A	PPB	27 Nov 2024 16:52:37	U	SMPL
P4995-02MS - 1	001MS	16152	3.3922	-	7470A	PPB	27 Nov 2024 16:54:53	U	SMPL
P4995-02MSD - 1	001MSD	15974	3.3536	-	7470A	PPB	27 Nov 2024 16:57:16	U	SMPL
CCV64 - 1	CCV64	24742	5.2570	-	7470A	PPB	27 Nov 2024 16:59:41	U	SMPL
CCB64 - 1	CCB64	541	0.0033	-	7470A	PPB	27 Nov 2024 17:05:57	U	SMPL
PB165248TB - 1	PB165248TB	794	0.0582	-	7470A	PPB	27 Nov 2024 17:08:12	U	SMPL
PB165252TB - 1	PB165252TB	782	0.0556	-	7470A	PPB	27 Nov 2024 17:10:27	U	SMPL
PB165159TB - 1	PB165159TB	787	0.0567	-	7470A	PPB	27 Nov 2024 17:12:44	U	SMPL
P5000-08LX5 - 1		-15	-0.1174	-	7470A	PPB	27 Nov 2024 17:14:59	U	SMPL
P5000-08A - 1		19086	4.0292	-	7470A	PPB	27 Nov 2024 17:17:18	U	SMPL
P4995-02LX5 - 1		30	-0.1077	-	7470A	PPB	27 Nov 2024 17:19:38	U	SMPL
P4995-02A - 1		20077	4.2443	-	7470A	PPB	27 Nov 2024 17:21:56	U	SMPL
CCV65 - 1	CCV65	23631	5.0158	-	7470A	PPB	27 Nov 2024 17:24:12	U	SMPL
CCB65 - 1	CCB65	350	-0.0382	-	7470A	PPB	27 Nov 2024 17:26:30	U	SMPL