

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB133666

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/27/2024	14:32
2	CAL2	1	Water	NA	NA	20.2	7.00	11/27/2024	14:33
3	CAL3	1	Water	NA	NA	20.2	10.02	11/27/2024	14:35
4	ICV	1	Water	NA	NA	20.2	7.00	11/27/2024	14:40
5	CCV1	1	Water	NA	NA	20.3	2.01	11/27/2024	14:45
6	P5005-02	1	Solid	20.02	20	24.7	8.87	11/27/2024	14:55
7	P5005-02DUP	1	Solid	20.03	20	24.8	8.88	11/27/2024	14:56
8	P5025-02	1	Solid	20.02	20	23.4	10.92	11/27/2024	15:10
9	P5025-06	1	Solid	20.03	20	23.1	10.14	11/27/2024	15:20
10	P5026-02	1	Solid	20.02	20	23.2	11.39	11/27/2024	15:27
11	P5026-06	1	Solid	20.03	20	22.9	10.04	11/27/2024	15:35
12	P5048-04	1	Solid	20.02	20	22.2	8.22	11/27/2024	15:44
13	CCV2	1	Water	NA	NA	20.3	12.02	11/27/2024	16:00