


LB133670

SFAM01.1

INSTRUMENT ID: CVJ



Linear

A= 0.0000e+000

B= 2.4111e-004 slope

C= -7.9660e-003 y-intercept

Rho= 0.9997667

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/20
0.00	0.000	0.016	0.016	100	0.000	100	0				
0.05	0.050					442	0				
0.20	0.200	0.206	0.006	887	0.0 %	887					
2.50	2.500	2.541	0.041	10570	0.0 %	10570					
5.00	5.000	4.847	-0.153	20137	0.0 %	20137					
7.50	7.500	7.611	0.111	31601	0.0 %	31601					
10.0	10.000	9.979	-0.021	41420	0.0 %	41420					

9/20
11
3
2
3
1
0

LB133670

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	100	-		0 SFAM01.1	PPB	11/27/2024 18:20	Std
	0.2 S01	887	-		0.2 SFAM01.1	PPB	11/27/2024 18:23	Std
	2.5 S02	10570	-		2.5 SFAM01.1	PPB	11/27/2024 18:25	Std
	5 S03	20137	-		5 SFAM01.1	PPB	11/27/2024 18:27	Std
	7.5 S04	31601	-		7.5 SFAM01.1	PPB	11/27/2024 18:30	Std
	10 S05	41420	-		10 SFAM01.1	PPB	11/27/2024 18:32	Std
ICV002	ICV002	15580	3.7485 -		SFAM01.1	PPB	11/27/2024 18:35	SMPL
ICB002	ICB002	-106	-0.0335 -		SFAM01.1	PPB	11/27/2024 18:37	SMPL
CCV066	CCV066	20002	4.8147 -		SFAM01.1	PPB	11/27/2024 18:39	SMPL
CCB066	CCB066	-71	-0.0251 -		SFAM01.1	PPB	11/27/2024 18:41	SMPL
PB165307BL	PBS3087	98	0.0157 -		SFAM01.1	PPB	11/27/2024 18:46	SMPL
P4879-01	E29X1	1306	0.3069 -		SFAM01.1	PPB	11/27/2024 18:49	SMPL
P4879-02	E29X2	858	0.1989 -		SFAM01.1	PPB	11/27/2024 18:51	SMPL
P4879-03	E29X3	894	0.2076 -		SFAM01.1	PPB	11/27/2024 18:53	SMPL
P4879-04	E29X4	590	0.1343 -		SFAM01.1	PPB	11/27/2024 18:55	SMPL
P4879-05	E29X5	947	0.2204 -		SFAM01.1	PPB	11/27/2024 18:58	SMPL
P4879-06	E29X5D	941	0.2189 -		SFAM01.1	PPB	11/27/2024 19:00	SMPL
P4879-07	E29X5S	12266	2.9495 -		SFAM01.1	PPB	11/27/2024 19:02	SMPL
P4879-08	E29X6	528	0.1193 -		SFAM01.1	PPB	11/27/2024 19:05	SMPL
P4879-09	E29X7	74	0.0099 -		SFAM01.1	PPB	11/27/2024 19:07	SMPL
P4879-10	E29X8	166	0.0321 -		SFAM01.1	PPB	11/27/2024 19:09	SMPL
P4879-11	E29X9	363	0.0796 -		SFAM01.1	PPB	11/27/2024 19:11	SMPL
P4879-12	E29Y0	251	0.0526 -		SFAM01.1	PPB	11/27/2024 19:14	SMPL
P4879-13	E29Y1	147	0.0275 -		SFAM01.1	PPB	11/27/2024 19:16	SMPL
P4879-14	E29Y2	123	0.0217 -		SFAM01.1	PPB	11/27/2024 19:18	SMPL
P4879-15	E29Y3	220	0.0451 -		SFAM01.1	PPB	11/27/2024 19:20	SMPL
P4879-16	E29Y4	96	0.0152 -		SFAM01.1	PPB	11/27/2024 19:23	SMPL
P4879-17	E29Y5	90	0.0137 -		SFAM01.1	PPB	11/27/2024 19:25	SMPL
P4879-18	E29Y9	40	0.0017 -		SFAM01.1	PPB	11/27/2024 19:27	SMPL
P4879-19	E29Z0	112	0.019 -		SFAM01.1	PPB	11/27/2024 19:30	SMPL
CCV067	CCV067	21195	5.1024 -		SFAM01.1	PPB	11/27/2024 19:32	SMPL
CCB067	CCB067	-58	-0.022 -		SFAM01.1	PPB	11/27/2024 19:34	SMPL
P4879-20	E29Y6	141	0.026 -		SFAM01.1	PPB	11/27/2024 19:36	SMPL
P4879-21	E29Y7	517	0.1167 -		SFAM01.1	PPB	11/27/2024 19:39	SMPL
P4879-22	E2A11	139	0.0255 -		SFAM01.1	PPB	11/27/2024 19:41	SMPL
PB165308BL	PBS308	102	0.0166 -		SFAM01.1	PPB	11/27/2024 19:43	SMPL
P4902-01	E29Y8	423	0.094 -		SFAM01.1	PPB	11/27/2024 19:45	SMPL
P4902-02	E2A12	431	0.096 -		SFAM01.1	PPB	11/27/2024 19:48	SMPL
P4902-03	E2A12D	332	0.0721 -		SFAM01.1	PPB	11/27/2024 19:50	SMPL
P4902-04	E2A12S	12106	2.9109 -		SFAM01.1	PPB	11/27/2024 19:52	SMPL
CCV068	CCV068	21094	5.078 -		SFAM01.1	PPB	11/27/2024 19:54	SMPL
CCB068	CCB068	-80	-0.0273 -		SFAM01.1	PPB	11/27/2024 19:57	SMPL