



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/3/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:50
In Date: 12/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 12/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133677

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5034-01	MBHKB4	1	1.18	8.53	9.71	8.55	86.4	
P5034-02	MBHKB5	2	1.18	8.77	9.95	8.56	84.2	
P5034-03	MBHKB6	3	1.17	8.40	9.57	8.27	84.5	
P5034-04	MBHKB6D	4	1.17	8.40	9.57	8.27	84.5	
P5034-05	MBHKB6S	5	1.17	8.40	9.57	8.27	84.5	
P5034-06	MBHKB7	6	1.19	8.72	9.91	8.84	87.7	
P5034-07	MBHKBH4	7	1.15	8.78	9.93	7.8	75.7	
P5034-08	MBHKBH5	8	1.19	8.77	9.96	8.37	81.9	
P5034-09	MBHKBH6	9	1.17	8.58	9.75	8.15	81.4	
P5034-10	MBHKBH7	10	1.15	8.68	9.83	8.02	79.1	
P5034-11	MBHKBH8	11	1.18	8.78	9.96	8.76	86.3	
P5034-12	MBHKBH9	12	1.15	8.84	9.99	9.24	91.5	
P5034-13	MBHKBKJ0	13	1.16	8.74	9.9	9.03	90.0	
P5034-14	MBHKBKP0	14	1.14	8.54	9.68	7.82	78.2	
P5034-15	MBHKBKP1	15	1.16	8.49	9.65	8.15	82.3	
P5034-16	MBHKBKP2	16	1.14	8.53	9.67	8.41	85.2	
P5034-17	MBHKBKP3	17	1.19	8.61	9.8	8.31	82.7	
P5034-18	MBHKBKP4	18	1.18	8.46	9.64	8.37	85.0	
P5034-19	MBHKBKP5	19	1.19	8.51	9.7	8.82	89.7	
P5034-20	MBHKBKP6	20	1.15	8.82	9.97	9.27	92.1	
P5034-21	MBHKBKP8	21	1.18	8.57	9.75	8.92	90.3	
P5034-22	MBHKBKP9	22	1.14	8.40	9.54	8.6	88.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$