



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/3/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:50
In Date: 12/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 12/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB133681

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P5036-01	MBHK33	1	1.18	8.59	9.77	8.49	85.1	
P5036-02	MBHKF3	2	1.19	8.63	9.82	7.52	73.3	
P5036-03	MBHKF4	3	1.15	8.82	9.97	7.94	77.0	
P5036-04	MBHKF5	4	1.17	8.60	9.77	8.13	80.9	
P5036-05	MBHKF6	5	1.19	8.58	9.77	8.52	85.4	
P5036-06	MBHKF7	6	1.16	8.44	9.6	8.58	87.9	
P5036-07	MBHKF8	7	1.19	8.73	9.92	9.29	92.8	
P5036-08	MBHKF8D	8	1.19	8.73	9.92	9.29	92.8	
P5036-09	MBHKF8S	9	1.19	8.73	9.92	9.29	92.8	
P5036-10	MBHKF9	10	1.14	8.85	9.99	8.83	86.9	
P5036-11	MBHKG0	11	1.15	8.83	9.98	7.91	76.6	
P5036-12	MBHKG1	12	1.18	8.59	9.77	8.34	83.4	
P5036-13	MBHKD5	13	1.15	8.53	9.68	8.11	81.6	
P5036-14	MBHKD7	14	1.17	8.80	9.97	8.85	87.3	
P5036-15	MBHKD8	15	1.19	8.67	9.86	8.84	88.2	
P5036-16	MBHKD9	16	1.12	8.86	9.98	8.04	78.1	
P5036-17	MBHKE0	17	1.15	8.40	9.55	7.93	80.7	
P5036-18	MBHKE1	18	1.18	8.67	9.85	8.36	82.8	
P5036-19	MBHKE2	19	1.15	8.36	9.51	8.3	85.5	
P5036-20	MBHKE3	20	1.16	8.46	9.62	8.68	88.9	
P5036-21	MBHKE4	21	1.13	8.76	9.89	8.88	88.5	
P5036-22	MBHK51	22	1.15	8.70	9.85	8.66	86.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$